ISO 13004:2022-10 (E)

Sterilization of health care products - Radiation - Substantiation of selected sterilization dose: Method VD<(Index)max><(hoch)SD>

Contents

Page

Foreword v

Introduction vi

1 Scope 1

2 Normative references 1

3 Termsanddefinitions 1

4 Definitionandmaintenanceofproductfamiliesforsterilizationdosesubstantiation and sterilization dose auditing 5
4.1 General 5
4.2 Defining product families 5
4.3 Designation of product to represent a product family 6
4.3.1 Product to represent a product family 6
4.3.2 Master product 7
4.3.3 Equivalent product 7
4.3.4 Simulated product 7
4.4 Maintaining product families 8
4.4.1 Periodic review 8
4.4.2 Modification to either product or manufacturing process, or both 8
4.4.3 Records 8
4.5 Consequence of failure of sterilization dose substantiation or sterilization dose audit 8

5 Selection and testing of product for substantiating and auditing a selected sterilization dose 8
5.1 Nature of product 8
5.2 Sample item portion (SIP) 9
5.3 Manner of sampling 10
5.4 Microbiological testing 11
5.5 Irradiation 11

6 Method VD SD max -- Substantiation of a selected sterilization dose of 17,5 kGy, 20 kGy, 22,5 kGy, 27,5 kGy, 30 kGy, 32,5 kGy or 35 kGy 12
6.1 Rationale 12
6.2 Procedure for Method VDmax SD for multiple production batches 12
6.2.1 General 12
6.2.2 Stage 1: Obtain samples of product 13
6.2.3 Stage 2: Determine average bioburden 13
6.2.4 Stage 3: Obtain the selected sterilization dose 13
6.2.5 Stage 4: Obtain VD SD max 14
6.2.6 Stage 5: Perform verification dose experiment 15
6.2.7 Stage 6: Interpretation of results 15
6.2.8 Confirmatory verification dose experiment 16
6.3 Procedure for Method VDmax SD for a single production batch 17
6.3.1 Rationale 17
6.3.2 General 17
6.3.3 Stage 1: Obtain samples of product 17
6.3.4 Stage 2: Determine average bioburden 18
6.3.5 Stage 3: Obtain the selected sterilization dose 18

6.3.6 Stage 4: Obtain VDmax SD 19
6.3.7 Stage 5: Perform verification dose experiment 19
6.3.8 Stage 6: Interpretation of results 19
6.3.9 Confirmatory verification dose experiment 20

7 Maintaining process effectiveness 21
7.1 General 21
7.2 Frequency of determination of bioburden 21
7.3 Sterilization dose audit 21
7.3.1 Frequency 21
7.3.2 Outcome 22
7.3.3 Procedure for auditing a sterilization dose substantiated using Method VDmax SD 22
7.3.4 Failure of a sterilization dose audit 25

8 Tables of values for SIP 26

9 Worked examples 51
9.1 Substantiation of a selected sterilization dose of 17,5 kGy (SIP less than 1,0) 51
9.2 Substantiation of a selected sterilization dose of 30 kGy (SIP equal to 1,0) 52
9.3 Sterilization dose audit for a sterilization dose substantiated using 52
9.4 Method VDmax 22,5 52

Bibliography 54