DIN EN ISO 13004:2023-10 (E)

Sterilization of health care products - Radiation - Substantiation of selected sterilization dose: Method VD_{max}^{SD} (ISO 13004:2022)

Contents

Page

European foreword ... **4**

Foreword ... **5**

Introduction ... **6**

1 Scope ... **7**

2 Normative references ... **7**

3 Terms and definitions ... **7**

4 Definition and maintenance of product families for sterilization dose substantiation and sterilization dose auditing ... **11**

4.1 General ... 11

4.2 Defining product families ... 11

4.3 Designation of product to represent a product family ... 12

4.3.1 Product to represent a product family ... 12

4.3.2 Master product ... 13

4.3.3 Equivalent product ... 13

4.3.4 Simulated product ... 13

4.4 Maintaining product families ... 14

4.4.1 Periodic review ... 14

4.4.2 Modification to either product or manufacturing process, or both ... 14

4.4.3 Records ... 14

4.5 Consequence of failure of sterilization dose substantiation or sterilization dose audit ... 14

5 Selection and testing of product for substantiating and auditing a selected sterilization dose ... **14**

5.1 Nature of product ... 14

5.2 Sample item portion (SIP) ... 15

5.3 Manner of sampling ... 16

5.4 Microbiological testing ... 17

5.5 Irradiation ... 17

6 Method VD_{max}^{SD} — Substantiation of a selected sterilization dose of 17,5 kGy, 20 kGy, 22,5 kGy, 27,5 kGy, 30 kGy, 32,5 kGy or 35 kGy ... **18**

6.1 Rationale ... 18

6.2 Procedure for Method VD_{max}^{SD} for multiple production batches ... 18

6.2.1 General ... 18

6.2.2 Stage 1: Obtain samples of product ... 19

6.2.3 Stage 2: Determine average bioburden ... 19

6.2.4 Stage 3: Obtain the selected sterilization dose ... 19

6.2.5 Stage 4: Obtain VD_{max}^{SD} ... 20

6.2.6 Stage 5: Perform verification dose experiment ... 21

6.2.7 Stage 6: Interpretation of results ... 21

6.2.8 Confirmatory verification dose experiment ... 22

6.3 Procedure for Method VD_{max}^{SD} for a single production batch ... 23

6.3.1 Rationale ... 23

6.3.2 General ... 23

6.3.3 Stage 1: Obtain samples of product ... 23

6.3.4 Stage 2: Determine average bioburden 24
6.3.5 Stage 3: Obtain the selected sterilization dose 24
6.3.6 Stage 4: Obtain VD_{max}^{SD} 25
6.3.7 Stage 5: Perform verification dose experiment 25
6.3.8 Stage 6: Interpretation of results 25
6.3.9 Confirmatory verification dose experiment 26

7 Maintaining process effectiveness 27
7.1 General 27
7.2 Frequency of determination of bioburden 27
7.3 Sterilization dose audit 27
7.3.1 Frequency 27
7.3.2 Outcome 28
7.3.3 Procedure for auditing a sterilization dose substantiated using Method VD_{max}^{SD} 28
7.3.4 Failure of a sterilization dose audit 31

8 Tables of values for SIP 32

9 Worked examples 57
9.1 Substantiation of a selected sterilization dose of 17,5 kGy (SIP less than 1,0) 57
9.2 Substantiation of a selected sterilization dose of 30 kGy (SIP equal to 1,0) 58
9.3 Sterilization dose audit for a sterilization dose substantiated using 58
9.4 Method $VD_{max}^{22,5}$ 58

Bibliography 60