

DIN EN ISO 13004:2023-10 (E)

Sterilization of health care products - Radiation - Substantiation of selected sterilization dose: Method VD_{max}^{SD} (ISO 13004:2022)

| Contents | Page |
|---|-----------|
| European foreword..... | 4 |
| Foreword..... | 5 |
| Introduction..... | 6 |
| 1 Scope..... | 7 |
| 2 Normative references..... | 7 |
| 3 Terms and definitions..... | 7 |
| 4 Definition and maintenance of product families for sterilization dose substantiation and sterilization dose auditing..... | 11 |
| 4.1 General..... | 11 |
| 4.2 Defining product families..... | 11 |
| 4.3 Designation of product to represent a product family..... | 12 |
| 4.3.1 Product to represent a product family..... | 12 |
| 4.3.2 Master product..... | 13 |
| 4.3.3 Equivalent product..... | 13 |
| 4.3.4 Simulated product..... | 13 |
| 4.4 Maintaining product families..... | 14 |
| 4.4.1 Periodic review..... | 14 |
| 4.4.2 Modification to either product or manufacturing process, or both..... | 14 |
| 4.4.3 Records..... | 14 |
| 4.5 Consequence of failure of sterilization dose substantiation or sterilization dose audit..... | 14 |
| 5 Selection and testing of product for substantiating and auditing a selected sterilization dose..... | 14 |
| 5.1 Nature of product..... | 14 |
| 5.2 Sample item portion (SIP)..... | 15 |
| 5.3 Manner of sampling..... | 16 |
| 5.4 Microbiological testing..... | 17 |
| 5.5 Irradiation..... | 17 |
| 6 Method VD_{max}^{SD} — Substantiation of a selected sterilization dose of 17,5 kGy, 20 kGy, 22,5 kGy, 27,5 kGy, 30 kGy, 32,5 kGy or 35 kGy..... | 18 |
| 6.1 Rationale..... | 18 |
| 6.2 Procedure for Method VD_{max}^{SD} for multiple production batches..... | 18 |
| 6.2.1 General..... | 18 |
| 6.2.2 Stage 1: Obtain samples of product..... | 19 |
| 6.2.3 Stage 2: Determine average bioburden..... | 19 |
| 6.2.4 Stage 3: Obtain the selected sterilization dose..... | 19 |
| 6.2.5 Stage 4: Obtain VD_{max}^{SD} | 20 |
| 6.2.6 Stage 5: Perform verification dose experiment..... | 21 |
| 6.2.7 Stage 6: Interpretation of results..... | 21 |
| 6.2.8 Confirmatory verification dose experiment..... | 22 |
| 6.3 Procedure for Method VD_{max}^{SD} for a single production batch..... | 23 |
| 6.3.1 Rationale..... | 23 |
| 6.3.2 General..... | 23 |
| 6.3.3 Stage 1: Obtain samples of product..... | 23 |

| | | |
|----------|---|-----------|
| 6.3.4 | Stage 2: Determine average bioburden | 24 |
| 6.3.5 | Stage 3: Obtain the selected sterilization dose | 24 |
| 6.3.6 | Stage 4: Obtain VD_{max}^{SD} | 25 |
| 6.3.7 | Stage 5: Perform verification dose experiment | 25 |
| 6.3.8 | Stage 6: Interpretation of results | 25 |
| 6.3.9 | Confirmatory verification dose experiment | 26 |
| 7 | Maintaining process effectiveness | 27 |
| 7.1 | General | 27 |
| 7.2 | Frequency of determination of bioburden | 27 |
| 7.3 | Sterilization dose audit | 27 |
| 7.3.1 | Frequency | 27 |
| 7.3.2 | Outcome | 28 |
| 7.3.3 | Procedure for auditing a sterilization dose substantiated using Method VD_{max}^{SD} | 28 |
| 7.3.4 | Failure of a sterilization dose audit | 31 |
| 8 | Tables of values for SIP | 32 |
| 9 | Worked examples | 57 |
| 9.1 | Substantiation of a selected sterilization dose of 17,5 kGy (SIP less than 1,0) | 57 |
| 9.2 | Substantiation of a selected sterilization dose of 30 kGy (SIP equal to 1,0) | 58 |
| 9.3 | Sterilization dose audit for a sterilization dose substantiated using | 58 |
| 9.4 | Method $VD_{max}^{22,5}$ | 58 |
| | Bibliography | 60 |