ISO 10555-4:2023-11 (E)

Intravascular catheters - Sterile and single-use catheters - Part 4: Balloon dilatation catheters

Contents

Page

Foreword iv

1 Scope 1

2 Normative references 1

3 Termsanddefinitions 1

4 Requirements 2
4.1 General 2
4.2 Detectability of the balloon position 2
4.3 Designation of nominal size 2
4.4 Physical requirements 2
4.4.1 Balloon rated burst pressure (RBP) 2
4.4.2 Balloon fatigue; freedom from leakage and damage on inflation 2
4.4.3 Balloon deflation time 2
4.4.4 Balloon diameter to inflation pressure (balloon compliance) 2
4.4.5 Crossing profile 3
4.4.6 Balloon removal 3
4.5 Information to be supplied with the catheter 3

Annex A (normative) Test for rated burst pressure (RBP) 4

Annex B (normative) Balloon fatigue test for freedom from leakage and damage oninflation 6

Annex C (normative)Testforballoondeflationtime 8

Annex D (normative)Testforballoondiametertoinflationpressure(balloooncompliance) 10

Annex E (normative)Determinationofcrossingprofile 12

Annex F (normative) Test method for balloon removal 14

Annex G (informative) Rationale and guidance 16

Bibliography 17