DIN EN 13060:2025-10 (E)

Sterilizers for medical purposes - Small steam sterilizers - Requirements and testing

Contents

Page

European foreword 6

Introduction 8

1 Scope 9

2 Normative references 9

3 Terms and definitions 11

4 General 23
4.1 Product definition 23
4.2 Equipment development 24
4.3 Calibration 24

5 Equipment design and construction 25
5.1 Safety and security 25
5.1.1 General requirements 25
5.1.2 Electromagnetic interference 26
5.1.3 Risk control and usability 26
5.2 Chamber 27
5.2.1 Dimensions 27
5.2.2 Doors 27
5.2.3 Chamber integrity 28
5.2.4 Pressure vessels 28
5.2.5 Uniformity of conditions 28
5.2.6 Ancillary equipment and components 28
5.3 Materials 29
5.4 Interlocks 29
5.5 Test connections 30
5.6 Vibration 30
5.7 User interfaces 31
5.8 Transport 31

6 Indicating, monitoring, controlling and recording 32
6.1 General 32
6.2 Automatic control 32
6.3 Software 33
6.4 Control and monitoring system 34
6.5 Failure 35
6.5.1 General 35
6.5.2 Fault 36
6.5.3 Power failure 36
6.5.4 Other failures 36
6.6 Instrumentation 37
6.7 Indicating devices 38
6.7.1 General 38
6.7.2 Sterilizer chamber temperature indicating instrument 38
6.7.3 Sterilizer chamber pressure instrument 38
6.7.4 Jacket pressure indicating instrument 39
6.7.5 Loading side of the sterilizer 39
6.7.6 Double ended sterilizer 40

6.7.7 Cycle counter 40
6.7.8 Air leak indication 41
6.8 Recorders 41
6.8.1 General 41
6.8.2 Recorders producing analogue records 42
6.8.3 Recorders producing digital records 43
6.9 Process 46
6.9.1 General 46
6.9.2 Sterilization temperature, sterilization temperature band, holding time 46
6.9.3 Time-temperature relationships 46
6.9.4 Equilibration time 46

7 Services and local environment 47
7.1 General 47
7.2 Sterilizing agent and sterilant 47
7.3 Electrical supply 47
7.4 Water 47
7.4.1 General 47
7.4.2 Water supply for steam generation in the sterilizer 47
7.4.3 Water used other than for steam generation 48
7.5 Steam 48
7.5.1 External steam supply 48
7.6 Vacuum 48
7.7 Drains 48
7.8 Lighting 49
7.9 Compressed air 49
7.10 Air and inert gases 49
7.11 Ventilation 49

8 Emissions 49
8.1 Electromagnetic emissions 49
8.2 Noise 49
8.3 Heat emission 50

9 Test instrumentation 50

10 Performance and assessment 51
10.1 General 51
10.2 Chamber integrity 52
10.2.1 General 52
10.2.2 Air leakage rate 52
10.3 Attainment of conditions 52
10.4 Performance verification 53
10.5 Drying 53
10.6 Microbiological performance 53
10.7 Pressure change 53
10.7.1 General 53
10.7.2 Dynamic sterilizer chamber pressure test 54

11 Information supplied by the manufacturer 54
11.1 General 54
11.2 Information to be made available prior to purchase 54
11.3 Marking 59
11.3.1 Marking of the pressure vessel 59
11.3.2 Marking of the sterilizer and the packaging 59
11.4 Label 60
11.5 Instructions for use 60
11.6 Technical description 61

Annex A (normative) Test programme 62

A.1 Categories of tests 62

A.2 Type tests 62
A.3 Works test 63
A.4 Installation tests 63
A.5 Test programme 63
A.6 Rationale for the tests (informative) 65

Annex B (normative) Test equipment 68

B.1 Test equipment 68
B.2 Temperature sensors 68
B.3 Thermometric recording instrument 68
B.4 Pressure measurement and recording instrument 69
B.5 Test equipment for the performance of the air leakage test 70
B.6 Process challenge device (PCD) and chemical indicator for narrow lumen 71
B.7 Process challenge device and chemical indicator for simple hollow item 71
B.8 Balance for load dryness test 72

Annex C (normative) Test loads 73

C.1 Test loads 73
C.2 Porous load 73
C.3 Solid load, unwrapped 75
C.4 Solid load, single wrapped 75
C.5 Solid load, double wrapped 75

Annex D (normative) Test procedures 76

D.1 Test procedures 76
D.2 Air leakage test 77
D.3 Dynamic sterilizer chamber pressure test 78
D.4 Empty chamber test 79
D.5 Solid load test 79
D.6 Narrow lumen test 79
D.7 Simple hollow item test 80
D.8 Small porous load test 81
D.9 Full porous load test (single and double wrapped) 82
D.10 Small porous items test (single and double wrapped) 83
D.11 Solid load dryness test 84
D.12 Porous load dryness test (small and full, single and double wrapped) 84
D.13 Small porous items dryness test (single and double wrapped) 85
D.14 Microbiological test for solid loads 85
D.15 Microbiological test for narrow lumens 86
D.16 Microbiological test for simple hollow item 86
D.17 Microbiological test for small porous loads 86
D.18 Microbiological test for full porous loads 87
D.19 Microbiological test for small porous items 87

Annex E (informative) Monitoring system 88

E.1 General 88
E.2 Principle 88
E.3 Equipment 89

Annex F (informative) Suggested maximum limits of contaminants in and specifications for water for steam sterilization 92

F.1 Suggested maximum limits of contaminants in and specifications for water for steam sterilization 92

Annex G (informative) Load support systems 93

Annex H (informative) Illustrations of the interrelationship between control and recording 94

H.1 Overview 94
H.2 Illustration 1 94
H.3 Illustration 2 96
H.4 Illustration 3 97

Annex I (informative) Environmental aspects 100

I.1 Environmental impact 100
I.2 Steam 100
I.3 Product life cycle 101

Annex ZA (informative) Relationship between this European Standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered 102

Bibliography 120