# ISO 7376-2:2025-09 (E)

## Anaesthetic and respiratory equipment - Part 2: Video laryngoscopes

## Contents

Page


Foreword ..... iv

Introduction ..... v

1 Scope ..... 1

2 Normative references ..... 1

3 Terms and definitions ..... 2

4 General requirements ..... 3
4.1 General ..... 3
4.2 Risk management ..... 3

5 Materials ..... 4
5.1 General ..... 4
5.2 Exterior / External surfaces ..... 4

6 Design requirements ..... 4
6.1 General ..... 4
6.2 Components ..... 4
6.3 Mechanical requirements ..... 5
6.3.1 VL strength ..... 5
6.3.2 Rigidity ..... 6
6.3.3 Drop test ..... 6
6.3.4 Illumination ..... 6
6.4 Image quality ..... 7
6.4.1 Limiting resolution ..... 7
6.4.2 Distortion ..... 7
6.4.3 Latency ..... 7
6.4.4 Contrast ..... 7
6.4.5 Image display ..... 7
6.4.6 Field of view (FOV) ..... 8
6.4.7 Camera lens condensation ..... 8
6.5 Temperature ..... 9
6.6 System communication ..... 9
6.7 Requirements for VLs for emergency medical services ..... 9
6.8 Transportation and storage ..... 9
6.9 Batteries ..... 9

7 Cleaning, disinfection and sterilization ..... 9

8 Information supplied by the manufacturer ..... 10
8.1 General ..... 10
8.2 Marking ..... 10
8.3 Reusable ..... 10
8.4 Instructions for use (IFU) ..... 10

Annex A (informative) Rationale ..... 12

Annex B (normative) Test method for illumination, strength and rigidity ..... 15

**Annex C (normative) Test method for determining the field of view ..... 17**

**Annex D (informative)Hazardidentificationforrisk assessment ..... 18**

**Annex E (informative) Video laryngoscope designs ..... 20**

**Bibliography ..... 22**