DIN EN ISO 17510:2026-07 (E)

Medical devices - Sleep apnoea breathing therapy - Masks and application accessories (ISO 17510:2025)

Contents

Page

Foreword ... **iv**

Introduction ... **v**

1 Scope ... **1**

2 Normative references ... **1**

3 Terms and definitions ... **2**

4 *Information supplied by the manufacturer* ... **8**
4.1 General ... 8
4.2 *Accompanying information* ... 8

5 Construction requirements ... **10**
5.1 *Mask* connectors ... 10
5.2 *Biocompatibility* ... 10
5.2.1 *Patient* contacting ... 10
5.2.2 *Gas pathway* contacting ... 10
5.3 Protection against *rebreathing* ... 10
5.3.1 *Normal condition* protection ... 11
5.3.2 *Single fault condition protection* ... 11
5.4 *Cleaning*, *disinfection*, and *sterilization* ... 11
5.4.1 *Single patient multiple use* ... 11
5.4.2 *Multiple patient multiple use* ... 12
5.5 Breathing during *single fault condition* ... 12
5.6 *Breathing system filter* ... 13

6 Audible acoustic energy ... **13**

7 Measurement uncertainty ... **13**

Annex A (informative) **Particular guidance and rationale** ... **14**

Annex B (normative) ***Exhaust flow* test *procedure*** ... **18**

Annex C (normative) **Resistance to flow (pressure drop)** ... **20**

Annex D (normative) ***Anti-asphyxia valve* pressure testing** ... **22**

Annex E (normative) **Determination of the inspiratory and expiratory pressure drop under *single fault condition*** ... **24**

Annex F (normative) **Carbon dioxide *rebreathing*** ... **26**

Annex G (normative) **Audible acoustic energy** ... **29**

Annex H (informative) **Guide to *information supplied by the manufacturer*** ... **31**

Annex I (informative) **Reference to the IMDRF essential principles and labelling principles** ... **32**

Annex J (informative) **Terminology — Alphabetized index of defined terms** ... **34**

Bibliography ... **36**