

ISO 11608-1:2000-12 (E)

Pen-injectors for medical use - Part 1: Pen-injectors; Requirements and test methods

Contents		Page
Foreword		iv
Introduction		v
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	Symbols and abbreviations	3
5	General requirements	4
6	Test conditions	5
6.1	Standard atmosphere	5
6.2	Cool atmosphere	5
6.3	Hot atmosphere	5
7	Preconditioning of pen-injectors	6
7.1	Preconditioning in dry heat atmosphere	6
7.2	Preconditioning in cold storage atmosphere	6
7.3	Preconditioning in cyclical atmosphere	6
7.4	Preconditioning by free fall	6
7.5	Preconditioning by vibration for pen-injectors with electronic components	7
8	Reagent and apparatus	7
9	Determination of dose accuracy	7
9.1	Dose accuracy	7
9.2	Dose accuracy requirements	12
10	Freedom from defects	14
10.1	Defects after being subjected to cyclical preconditioning	14
10.2	Freedom from defects after being subjected to vibration	14
10.3	Freedom from defects after being subjected to free fall	14
11	Determination of electromagnetic compatibility	14
11.1	Electromagnetic compatibility (EMC)	14
11.2	Electrostatic discharge	15
11.3	Radiated radio frequency (RF) fields	15
12	Visual inspection	15
13	Functional inspection	16
13.1	Replaceable cartridge	16
13.2	Nonreplaceable cartridge	16
13.3	Accuracy	16
14	Test report	16
15	Information supplied by the manufacturer	16

15.1	General	16
15.2	Marking	17
15.3	Instructions for use	17
Annex A (informative) Two-sided tolerance limit factors (k)		19
Bibliography		25