ISO 11608-1:2000-12 (E)

Pen-injectors for medical use - Part 1: Pen-injectors; Requirements and test methods

Contents

Page

Foreword iv

Introduction v

1 Scope 1

2 Normative references 1

3 Terms and definitions 1

4 Symbols and abbreviations 3

5 General requirements 4

6 Test conditions 5
6.1 Standard atmosphere 5
6.2 Cool atmosphere 5
6.3 Hot atmosphere 5

7 Preconditioning of pen-injectors 6
7.1 Preconditioning in dry heat atmosphere 6
7.2 Preconditioning in cold storage atmosphere 6
7.3 Preconditioning in cyclical atmosphere 6
7.4 Preconditioning by free fall 6
7.5 Preconditioning by vibration for pen-injectors with electronic components 7

8 Reagent and apparatus 7

9 Determination of dose accuracy 7
9.1 Dose accuracy 7
9.2 Dose accuracy requirements 12

10 Freedom from defects 14
10.1 Defects after being subjected to cyclical preconditioning 14
10.2 Freedom from defects after being subjected to vibration 14
10.3 Freedom from defects after being subjected to free fall 14

11 Determination of electromagnetic compatibility 14
11.1 Electromagnetic compatibility (EMC) 14
11.2 Electrostatic discharge 15
11.3 Radiated radio frequency (RF) fields 15

12 Visual inspection 15

13 Functional inspection 16
13.1 Replaceable cartridge 16
13.2 Nonreplaceable cartridge 16
13.3 Accuracy 16

14 Test report 16

15 Information supplied by the manufacturer 16

15.1 General 16
15.2 Marking 17
15.3 Instructions for use 17

Annex A (informative) Two-sided tolerance limit factors (k) 19

Bibliography 25