

DIN ISO 15759:2006-05 (E)

Medical infusion equipment - Plastics caps with inserted elastomeric liner for containers manufactured by the blow-fill-seal (BFS) process (ISO 15759:2005)

| Contents | | Page |
|----------------------------|--|-------------|
| Foreword | | 2 |
| Introduction | | 2 |
| 1 | Scope | 3 |
| 2 | Normative references | 3 |
| 3 | Dimensions and designation | 4 |
| 3.1 | Plastics cap for attachment by welding technique (Form A) | 4 |
| 3.2 | Plastics cap for attachment by collar technique (Form B) | 5 |
| 3.3 | Elastomeric liner | 6 |
| 4 | Materials for cap and liner | 6 |
| 5 | Plastics cap -- Physical requirements and testing | 7 |
| 5.1 | Leak-resistance test | 7 |
| 5.2 | Opening force | 7 |
| 6 | Liner -- Physical requirements and testing | 7 |
| 6.1 | General requirements | 7 |
| 6.2 | Hardness | 7 |
| 6.3 | Fragmentation (coring) | 7 |
| 6.4 | Penetration force | 7 |
| 6.5 | Dynamic spike-retention capability | 7 |
| 6.6 | Static spike-retention capability of the liner and leak resistance of the piercing area | 7 |
| 6.7 | Resealability | 7 |
| 7 | Plastics cap -- Chemical requirements and testing | 7 |
| 8 | Liner -- Chemical requirements and testing | 8 |
| 9 | Biological requirements for plastics cap and liner | 8 |
| 10 | Packaging | 8 |
| 11 | Storage | 8 |
| 12 | Marking | 8 |
| Annex A (normative) | Leak-resistance test | 9 |
| Annex B (normative) | Opening force needed to expose the piercing area | 10 |
| Annex C (normative) | Fragmentation (coring) | 11 |
| Annex D (normative) | Penetration force | 12 |
| Annex E (normative) | Dynamic spike retention capability | 14 |

| | |
|--|-----------|
| Annex F (normative) Static spike-retention capability of the liner and leak resistance of the piercing area | 15 |
| Annex G (normative) Resealability | 16 |
| Annex H (normative) Holding device for elastomeric liner | 17 |
| Annex I (normative) Reference spike | 18 |
| Bibliography | 19 |