

DIN EN ISO 23747:2007-10 (E)

Anaesthetic and respiratory equipment - Peak expiratory flow meters for the assessment of pulmonary function in spontaneously breathing humans (ISO 23747:2007)

Contents

Page

Foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 General requirements	6
4.1 Safety for PEFMs that utilize electricity	6
4.2 Mechanical safety for all PEFMs	6
5 Identification, marking and documents.....	6
5.1 Marking of the scale or display	6
5.2 Marking of PEFM or packaging	7
5.3 Instructions for use	7
5.4 Technical description.....	8
6 PEFM measurement range.....	8
7 Performance requirements	8
7.1 Error of measurement	8
7.2 Linearity	8
7.3 Resistance to flow	9
7.4 Frequency response	9
8 Dismantling and reassembly	9
9 Effects of mechanical ageing	9
10 Effects of dropping a hand-held PEFM	9
11 Cleaning, sterilization and disinfection	9
11.1 Re-usable PEFM and parts	9
11.2 PEFM and parts delivered sterile	10
12 Compatibility with substances	10
13 Biocompatibility	10
Annex A (informative) Rationale for tests and examples of test apparatus	11
Annex B (normative) Determination of error, repeatability and resistance to PEFM output	14
Annex C (normative) Determination of frequency response	17
Annex D (normative) Test methods for determination of the effects of dismantling, ageing and dropping	19
Annex E (informative) Environmental aspects	21
Annex F (informative) Reference to the essential principals	23
Bibliography	26
Annex ZA (informative) Relationship between this International Standard and the Essential Requirements of EU Directive 93/42/EEC.....	28