ISO 11140-5:2007-03 (E)

Sterilization of health care products - Chemical indicators - Part 5: Class 2 indicators for Bowie and Dick-type air removal tests

Contents

Page

Foreword iv

Introduction v

1 Scope 1

2 Normative references 1

3 Terms and definitions 1

4 General requirements 2

5 Indicator system 2
5.1 Format 2
5.2 Performance 2

6 Indicator 3
6.1 Format 3
6.2 Performance 3

7 Packaging and labelling 4

8 Quality assurance 4

9 Sampling conditioning 4

Annex A (normative) Determination of the degree of contrast between the colour of the substrate and the indicator agent 5

Annex B (normative) Method of determining uniform colour change on exposure to saturated steam 6

Annex C (normative) Method for determining indicator colour change on exposure to dry heat 7

Annex D (normative) Method for determining transfer of indicator agent to standard test pack 8

Annex E (normative) Standard test pack 9

Annex F (normative) Method of determining non-uniform colour change on exposure to a standard fault condition 10

Annex G (normative) Steam exposure apparatus 11

Bibliography 12