DIN EN 16274:2012-12 (E)

Methods for analysis of allergens - Quantification of suspected fragrance allergens in consumer pro ducts - Step 1: GC analysis of ready-to-inject sample

Contents

Page

Foreword 3

Introduction 4

1 Scope 5

2 Principle 5

3 Reagents 5
3.1 Solvents 5
3.2 Fragrance (suspected allergen) standards 5
3.3 Internal standards (ISTD) 7

4 Apparatus 7
4.1 Analytical Balance 7
4.2 GC-FID (for solvent or standard purity only) 7
4.3 GC-MS 7
4.4 GC capillary columns 7

5 Standard preparation and preservation 8
5.1 General 8
5.2 Standard preparation 8
5.2.1 General 8
5.2.2 Stock solution of all allergens (5 g/l) 8
5.2.3 Separate stock solutions (carbonyl / non carbonyl compounds) (10 g/l) 9
5.2.4 Internal standard solution 9
5.2.5 Working solutions 9
5.2.6 Calibration solution 9

6 Procedure 10
6.1 General 10
6.2 Chromatographic conditions 10
6.3 MS conditions 10
6.3.1 General 10
6.3.2 SCAN mode 10
6.3.3 SIM mode 11
6.4 Calibration 13

7 Sample analysis 13

8 Data treatment and calculation of results 14
8.1 Identification of allergens 14
8.2 Quantification of allergens 14
8.3 Assessment of the analytical measurement 15
8.3.1 General 15
8.3.2 Examination of the Q-values 15
8.3.3 Maximum permitted tolerances 15

9 Test report 16

Annex A (informative) Column performances 17

Annex B (informative) SIM windows 19

Annex C (informative) Example of chromatographic separation 20

Annex D (normative) Decisional tree 22

Bibliography 23