

ISO 8041:2005-04 (E)

Human response to vibration - Measuring instrumentation

Contents		Page
Foreword		v
1	Scope	1
2	Normative references	1
3	Terms, definitions and symbols	2
3.1	Terms and definitions	2
3.2	Symbols	5
4	Reference environmental conditions	7
5	Performance specifications	7
5.1	General characteristics	7
5.2	Display of signal magnitude	9
5.3	Electrical output	10
5.4	Vibration sensitivity	10
5.5	Accuracy of indication at reference frequency under reference conditions	10
5.6	Frequency weightings and frequency responses	11
5.7	Amplitude linearity	14
5.8	Instrument noise	14
5.9	Signal-burst response	14
5.10	Overload indication	17
5.11	Under-range indication	18
5.12	Time averaging	18
5.13	Running r.m.s. acceleration	18
5.14	Reset	18
5.15	Timing facilities	19
5.16	Electrical cross-talk	19
5.17	Vibration transducer characteristics	19
5.18	Power supply	19
6	Mounting	19
7	Environmental and electromagnetic criteria	20
7.1	General	20
7.2	Air temperature	20
7.3	Surface temperature	20
7.4	Electrostatic discharge	20
7.5	Radio-frequency emissions and public-power-supply disturbances	21
7.6	Immunity to a.c. power-frequency fields and radio-frequency fields	21
7.7	Ingress of water and dust	22
8	Provision for use with auxiliary devices	22
9	Instrument marking	22
10	Instrument documentation	23
11	Testing and calibration	23
12	Pattern evaluation	24

12.1	Introduction	24
12.2	Testing requirements	25
12.3	Submission for testing	25
12.4	Marking of the vibration meter and information in the instrument documentation	25
12.5	Mandatory facilities and general requirements	26
12.6	Initial instrument preparation	26
12.7	Indication at the reference frequency under reference conditions	26
12.8	Electrical cross-talk	27
12.9	Vibration transducer	27
12.10	Amplitude linearity and under-range indication	27
12.11	Frequency weightings and frequency responses	29
12.12	Instrument noise	32
12.13	Signal-burst response	32
12.14	Overload indication	32
12.15	Reset	33
12.16	Combined axis outputs	33
12.17	A.c. electrical output	33
12.18	Timing facilities	33
12.19	Power supply	34
12.20	Environmental, electrostatic and radio-frequency tests	34
12.21	Test report	38
13	Verification tests	38
13.1	Introduction	38
13.2	Testing requirements	38
13.3	Submission for a test	39
13.4	Marking of the vibration meter and information in the instrument documentation	39
13.5	Mandatory facilities and general requirements	39
13.6	Initial instrument preparation	39
13.7	Indication at the reference frequency under reference conditions	39
13.8	Electrical cross-talk	40
13.9	Amplitude linearity and under-range indication	40
13.10	Frequency weightings and frequency responses	41
13.11	Instrument noise	43
13.12	Signal-burst response	43
13.13	Overload indication	43
13.14	Reset	44
13.15	Combined axis outputs	44
13.16	Test report	44
14	In-situ checks	44
14.1	Introduction	44
14.2	Preliminary inspection	45
14.3	Vibration sensitivity (field calibration)	45
Annex A (normative) Specification for vibration field calibrator		46
Annex B (informative) Frequency weightings		49
Annex C (informative) Realization of frequency weighting filters		68
Annex D (informative) Running r.m.s. time averaging		72
Annex E (informative) Vibration transducer characteristics		74
Annex F (informative) Tests for mounting systems		76
Annex G (normative) Instrument documentation		79
Annex H (normative) Phase-response requirements for measurement of non-r.m.s. quantities		84
Bibliography		91