

DIN EN ISO 20857:2013-08 (E)

Sterilization of health care products - Dry heat - Requirements for the development, validation and routine control of a sterilization process for medical devices (ISO 20857:2010)

Contents

Page

Foreword	4
Introduction	5
1 Scope	7
1.1 Inclusions	7
1.2 Exclusions	7
2 Normative references	8
3 Terms and definitions	8
4 Quality management system elements	16
4.1 Documentation	16
4.2 Management responsibility	16
4.3 Product realization	16
4.4 Measurement, analysis and improvement -- Control of nonconforming product	16
5 Sterilizing agent characterization	17
5.1 Sterilizing agent	17
5.2 Microbicidal effectiveness	17
5.3 Material effects	17
5.4 Environmental considerations	17
6 Process and equipment characterization	17
6.1 Process characterization	17
6.2 Equipment characterization	17
7 Product definition	19
7.1 General	19
7.2 Product safety and performance	19
7.3 Packaging considerations	20
7.4 Microbiological quality	20
7.5 Product family	20
7.6 Biological safety	20
8 Process definition	21
9 Validation	22
9.1 General	22
9.2 Installation qualification	22
9.3 Operational qualification	22
9.4 Performance qualification	22
9.5 Additional sterilization systems	24
9.6 Review and approval of validation	24
10 Routine monitoring and control	25
10.1 Routine control	25
10.2 Routine monitoring	25

10.3	Process monitoring locations	26
11	Product release from sterilization/depyrogenation	27
12	Maintaining process effectiveness	27
12.1	General	27
12.2	Recalibration	27
12.3	Maintenance of equipment	27
12.4	Requalification	27
12.5	Assessment of change	28
Annex A (informative) Guidance on the application of this International Standard		29
Annex B (informative) Process definition based on inactivation of the microbial population in its natural state (bioburden-based approach)		52
Annex C (informative) Process definition based on the inactivation of reference micro- organisms and knowledge of bioburden (combined bioburden/biological indicator approach)		54
Annex D (informative) Conservative process definition based on inactivation of reference microorganisms (overkill method)		57
Annex E (informative) Process development		60
Bibliography		63
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 90/385/EEC on Active Implantable Medical Devices		65
Annex ZB (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on Medical Devices		66
Annex ZC (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 98/79/EC on in vitro diagnostic medical devices		67