DIN 6868-157:2014-11 (E)

Image quality assurance in diagnostic X-ray departments - Part 157: X-ray ordinance acceptance and constancy test of image display systems in their environment

Contents

Page

Foreword ... 4

1 Scope ... 6

2 Normative references ... 6

3 Terms and definitions ... 7

4 Symbols and abbreviations ... 13

5 Test conditions ... 14

6 Test equipment ... 15
6.1 General ... 15
6.2 Luminance meter ... 15
6.2.1 Luminance meter for the acceptance test ... 15
6.2.2 Luminance meter for the constancy test ... 15
6.3 Illuminance meter / luxmeter ... 15
6.4 Test images ... 16

7 Room classes ... 17

8 Acceptance test for image display systems with diagnostic quality ... 19
8.1 Scope of testing ... 19
8.2 Visual tests ... 20
8.2.1 General ... 20
8.2.2 Overall image quality ... 20
8.2.3 Greyscale resolution of the IMAGE DISPLAY DEVICE ... 22
8.2.4 LUMINANCE uniformity (visual) ... 23
8.2.5 Colour impression and uniformity ... 23
8.2.6 Pixel errors ... 24
8.2.7 Image geometry ... 24
8.2.8 Test based on clinical reference images ... 25
8.2.9 Additional tests for mammography image display systems ... 25
8.3 Metrological tests ... 25
8.3.1 Minimum luminance ... 25
8.3.2 Maximum luminance ... 25
8.3.3 Ambient luminance ... 25
8.3.4 Illuminance ... 25
8.3.5 Maximum luminance ratio ... 25
8.3.6 Uniformity of the image display device ... 26
8.3.7 Uniformity of multiple image display devices ... 27
8.3.8 Luminance response curve ... 27
8.4 Minimum requirements (without mammography) ... 28
8.5 Additional requirements for image display systems used for mammography ... 30
8.6 Mobile systems ... 31

9 Reference values for the constancy test ... 31

10 Constancy test for image display systems with diagnostic quality ... 31

11 Significant changes to an image display system **33**

12 Image display systems with review quality **33**

Annex A (normative) **Methods of measuring luminance** **34**

A.1 General **34**

A.2 Measurement methods **34**
A.2.1 Method A: Telescopic method **34**
A.2.2 Method B: Close-range luminance meter in combination with a luxmeter **35**
A.2.3 Method C: Front-integrated luminance meter in combination with a luxmeter **36**
A.2.4 Method D: Back-integrated luminance meter in combination with an illuminance meter/luxmeter **37**
A.3 Notes on the measuring methods **37**

Annex B (informative) **Typical errors** **38**

Annex C (informative) **Test report templates** **39**
C.1 Example of a test report template for the acceptance test **39**
C.2 Example of a test report template for the constancy test **50**
C.3 Example of a test report template for the daily constancy test **52**

Bibliography **53**

Keyword index **55**

Figures

Figure 1 — Test images for use on image display systems (images a) through f) according to DIN EN 62563-1:2014-01, Table 2, and images g) through l) according to AAPM) **16**

Figure 2 — Overall image quality test using test image TG18-OIQ **22**

Figure 3 — Enlarged view of test image TG18-MP, depicting the 8-bit and 10-bit markers **23**

Figure 4 — Positions of the measuring points for the luminance uniformity test **26**

Figure A.1 — Method A – Telescopic method **34**

Figure A.2 — Method B – Close-range luminance meter in combination with a luxmeter **35**

Figure A.3 — Method C – Front-integrated luminance meter in combination with luxmeter **36**

Figure A.4 — Method D – Back-integrated luminance meter in combination with a luxmeter **37**

Tables

Table 1 — Definition of room classes with their respective requirements **18**

Table 2 — Visual tests for acceptance testing **19**

Table 3 — Metrological tests for acceptance testing **20**

Table 4 — Minimum requirements for image display devices **28**

Table 5 — Additional requirements for image display systems used in mammography and stereotactic mammography **30**

Table 6 — Visual tests for the constancy test **32**

Table 7 — Metrological test items in the constancy test **32**