

DIN EN IEC 63073-1:2021-09 (E)

Dedicated radionuclide imaging devices - Characteristics and test conditions - Part 1: Cardiac SPECT (IEC 63073-1:2020)

Contents

	Page
European foreword	3
Annex ZA (normative) Normative references to international publications with their corresponding European publications	4
INTRODUCTION.....	5
1 Scope.....	6
2 Normative references	6
3 Terms and definitions	6
4 Test methods.....	7
4.1 General.....	7
4.2 Detector characteristics	8
4.2.1 General	8
4.2.2 Energy resolution and LOW-ENERGY-TAIL RATIO measurement	8
4.2.3 Shield leakage.....	9
4.2.4 COUNT RATE performance	10
4.2.5 System sensitivity	12
4.2.6 Non-uniformity for each CARDIAC DETECTOR HEAD	14
4.2.7 SCATTER FRACTION	14
4.3 Characteristics of tomographic images.....	16
4.3.1 CENTRE OF ROTATION (COR)	16
4.3.2 REFERENCE POINT localization in the reconstructed FOV.....	16
4.3.3 Accuracy of tomographic system sensitivity modelling	17
4.3.4 Tomographic SPATIAL NON-LINEARITY.....	19
4.3.5 Tomographic SPATIAL RESOLUTION	21
4.3.6 Image quality assessment using a heart phantom	23
5 Additional testing	26
6 ACCOMPANYING DOCUMENTS	27
Bibliography.....	28
Index of defined terms	29
 Figure 1 – Small shielded liquid source.....	10
Figure 2 – Transverse slice of phantom used for measuring COUNT RATE performance	11
Figure 3 – Evaluation of SCATTER FRACTION	15
Figure 4 – Calculation of FWHM and measurement of the location of the maximum value.....	17
Figure 5 – Transaxial view of the 7 LINE SOURCE Phantom	18
Figure 6 – Transaxial view of the 7 LINE SOURCE phantom centred within a 140 mm diameter water-filled cylinder	22

Figure 7 – Placement of ROIs in SHORT AXIS view of myocardium.....	25
Figure 8 – Placement of ROIs in LONG AXIS view of myocardium.....	26
Table 1 – Relative ACTIVITY concentration of compartments of the anthropomorphic phantom	24