

# DIN EN IEC 63073-1:2021-09 (E)

## Dedicated radionuclide imaging devices - Characteristics and test conditions - Part 1: Cardiac SPECT (IEC 63073-1:2020)

---

Contents	Page
European foreword .....	3
Annex ZA (normative) Normative references to international publications with their corresponding European publications .....	4
INTRODUCTION.....	5
1 Scope.....	6
2 Normative references .....	6
3 Terms and definitions .....	6
4 Test methods.....	7
4.1 General.....	7
4.2 Detector characteristics .....	8
4.2.1 General .....	8
4.2.2 Energy resolution and LOW-ENERGY-TAIL RATIO measurement .....	8
4.2.3 Shield leakage.....	9
4.2.4 COUNT RATE performance.....	10
4.2.5 System sensitivity .....	12
4.2.6 Non-uniformity for each CARDIAC DETECTOR HEAD .....	14
4.2.7 SCATTER FRACTION .....	14
4.3 Characteristics of tomographic images.....	16
4.3.1 CENTRE OF ROTATION (COR) .....	16
4.3.2 REFERENCE POINT localization in the reconstructed FOV.....	16
4.3.3 Accuracy of tomographic system sensitivity modelling .....	17
4.3.4 Tomographic SPATIAL NON-LINEARITY.....	19
4.3.5 Tomographic SPATIAL RESOLUTION .....	21
4.3.6 Image quality assessment using a heart phantom.....	23
5 Additional testing.....	26
6 ACCOMPANYING DOCUMENTS .....	27
Bibliography.....	28
Index of defined terms .....	29
Figure 1 – Small shielded liquid source.....	10
Figure 2 – Transverse slice of phantom used for measuring COUNT RATE performance .....	11
Figure 3 – Evaluation of SCATTER FRACTION .....	15
Figure 4 – Calculation of FWHM and measurement of the location of the maximum value.....	17
Figure 5 – Transaxial view of the 7 LINE SOURCE Phantom .....	18
Figure 6 – Transaxial view of the 7 LINE SOURCE phantom centred within a 140 mm diameter water-filled cylinder .....	22

Figure 7 – Placement of ROIs in SHORT AXIS view of myocardium.....25

Figure 8 – Placement of ROIs in LONG AXIS view of myocardium.....26

Table 1 – Relative ACTIVITY concentration of compartments of the anthropomorphic phantom .....24