

DIN 6868-157:2022-01 (E)

Image quality assurance in diagnostic X-ray departments - Part 157: X-ray ordinance acceptance and constancy test of image display systems in their environment

Contents		Page
Foreword		4
1	Scope	7
2	Normative references	7
3	Terms and definitions	7
4	Symbols and abbreviated terms	13
4.1	Symbols	13
4.2	Abbreviated terms	14
5	Test conditions	14
6	Test equipment	15
6.1	General	15
6.2	Luminance meter	16
6.2.1	Luminance meter for the acceptance test	16
6.2.2	Luminance meter for the constancy test	16
6.3	Illuminance meter/luxmeter	16
6.4	Test images	16
7	Room classes	17
8	Acceptance test for image display systems with diagnostic quality	18
8.1	Scope of testing	18
8.2	Visual tests	20
8.2.1	General	20
8.2.2	Overall image quality	20
8.2.3	Greyscale resolution of the image display device	22
8.2.4	Luminance uniformity (visual)	23
8.2.5	Colour impression and uniformity	23
8.2.6	Pixel errors	23
8.2.7	Image geometry	24
8.2.8	Test based on clinical reference images	24
8.2.9	Additional tests for mammography image display systems	24
8.3	Metrological tests	24
8.3.1	Minimum luminance	24
8.3.2	Maximum luminance	24
8.3.3	Ambient luminance	24
8.3.4	Illuminance	24
8.3.5	Maximum luminance ratio	25
8.3.6	Uniformity of the image display device	25
8.3.7	Uniformity of multiple image display devices	26
8.3.8	Luminance response curve	26
8.4	Minimum requirements for image display systems (without mammography)	26
8.5	Minimum requirements for image display systems used in mammography and stereotactic mammography	28
8.6	Mobile systems	30
9	Reference values for the constancy test	30

10	Constancy test for image display systems with diagnostic quality	30
11	Significant changes to an image display system	31
12	Image display systems with review quality	32
Annex A (normative) Methods of measuring luminance		33
A.1	General	33
A.2	Measurement methods	33
A.2.1	Method A: Telescopic method	33
A.2.2	Method B: Close-range luminance meter in combination with an illuminance meter (luxmeter)	34
A.2.3	Method C: Front-integrated luminance meter in combination with an illuminance meter (luxmeter)	35
A.2.4	Method D: Back-integrated luminance meter in combination with an illuminance meter (luxmeter)	36
A.3	Notes on the measuring methods	36
Annex B (informative) Typical errors		37
Annex C (informative) Test report templates		38
C.1	Example of a test report template for the acceptance test	38
C.2	Example of a test report template for the constancy test	49
C.3	Example of a test report template for the daily constancy test	55
Bibliography		56
Index of defined terms		58
Figures Figure 1 -- Test images for use on image display systems (images a) to f) as in DIN EN 62563-1:2014-01, Table 4, and g) to l) as in AAPM)		17
Figure 2 -- Overall image quality test using test image TG18-OIQ		21
Figure 3 -- Enlarged view of test image TG18-MP, depicting the 8-bit and 10-bit markers		22
Figure 4 -- Positions of the measuring points for the luminance uniformity test		25
Figure A.1 -- Method A -- Telescopic method		33
Figure A.2 -- Method B --Close-range luminance meter in combination with an illuminance meter (luxmeter)		34
Figure A.3 -- Method C --Front-integrated luminance meter in combination with an illuminance meter (luxmeter)		35
Figure A.4 -- Method D -- Back-integrated luminance meter in combination with an illuminance meter (luxmeter)		36
Tables Table 1 -- Symbols		13
Table 2 -- Abbreviated terms		14
Table 3 -- Definition of room classes with their respective requirements		18
Table 4 -- Visual tests for acceptance testing		19
Table 5 -- Metrological tests for acceptance testing		19

Table 6 -- Minimum requirements for image display devices	27
Table 7 -- Minimum requirements for image display systems used in mammography and stereotactic mammography	29
Table 8 -- Visual tests for the constancy test	31
Table 9 -- Metrological test items in the constancy test	31