DIN 6868-157:2022-01 (E)

Image quality assurance in diagnostic X-ray departments - Part 157: X-ray ordinance acceptance and constancy test of image display systems in their environment

Contents

Page

Foreword 4

1 Scope 7

2 Normative references 7

3 Terms and definitions 7

4 Symbols and abbreviated terms 13
4.1 Symbols 13
4.2 Abbreviated terms 14

5 Test conditions 14

6 Test equipment 15
6.1 General 15
6.2 Luminance meter 16
6.2.1 Luminance meter for the acceptance test 16
6.2.2 Luminance meter for the constancy test 16
6.3 Illuminance meter/luxmeter 16
6.4 Test images 16

7 Room classes 17

8 Acceptance test for image display systems with diagnostic quality 18
8.1 Scope of testing 18
8.2 Visual tests 20
8.2.1 General 20
8.2.2 Overall image quality 20
8.2.3 Greyscale resolution of the image display device 22
8.2.4 Luminance uniformity (visual) 23
8.2.5 Colour impression and uniformity 23
8.2.6 Pixel errors 23
8.2.7 Image geometry 24
8.2.8 Test based on clinical reference images 24
8.2.9 Additional tests for mammography image display systems 24
8.3 Metrological tests 24
8.3.1 Minimum luminance 24
8.3.2 Maximum luminance 24
8.3.3 Ambient luminance 24
8.3.4 Illuminance 24
8.3.5 Maximum luminance ratio 25
8.3.6 Uniformity of the image display device 25
8.3.7 Uniformity of multiple image display devices 26
8.3.8 Luminance response curve 26
8.4 Minimum requirements for image display systems (without mammography) 26
8.5 Minimum requirements for image display systems used in mammography and stereotactic mammography 28
8.6 Mobile systems 30

9 Reference values for the constancy test 30

10 Constancy test for image display systems with diagnostic quality 30

11 Significant changes to an image display system 31

12 Image display systems with review quality 32

Annex A (normative) Methods of measuring luminance 33

A.1 General 33
A.2 Measurement methods 33
A.2.1 Method A: Telescopic method 33
A.2.2 Method B: Close-range luminance meter in combination with an illuminance meter (luxmeter) 34
A.2.3 Method C: Front-integrated luminance meter in combination with an illuminance meter (luxmeter) 35
A.2.4 Method D: Back-integrated luminance meter in combination with an illuminance meter (luxmeter) 36
A.3 Notes on the measuring methods 36

Annex B (informative) Typical errors 37

Annex C (informative) Test report templates 38

C.1 Example of a test report template for the acceptance test 38
C.2 Example of a test report template for the constancy test 49
C.3 Example of a test report template for the daily constancy test 55

Bibliography 56

Index of defined terms 58

Figures Figure 1 -- Test images for use on image display systems (images a) to f) as in DIN EN 62563-1:2014-01, Table 4, and g) to l) as in AAPM) 17

Figure 2 -- Overall image quality test using test image TG18-OIQ 21

Figure 3 -- Enlarged view of test image TG18-MP, depicting the 8-bit and 10-bit markers 22

Figure 4 -- Positions of the measuring points for the luminance uniformity test 25

Figure A.1 -- Method A -- Telescopic method 33

Figure A.2 -- Method B --Close-range luminance meter in combination with an illuminance meter (luxmeter) 34

Figure A.3 -- Method C --Front-integrated luminance meter in combination with an illuminance meter (luxmeter) 35

Figure A.4 -- Method D -- Back-integrated luminance meter in combination with an illuminance meter (luxmeter) 36

Tables Table 1 -- Symbols 13

Table 2 -- Abbreviated terms 14

Table 3 -- Definition of room classes with their respective requirements 18

Table 4 -- Visual tests for acceptance testing 19

Table 5 -- Metrological tests for acceptance testing 19

Table 6 -- Minimum requirements for image display devices 27

Table 7 -- Minimum requirements for image display systems used in mammography and stereotactic mammography 29

Table 8 -- Visual tests for the constancy test 31

Table 9 -- Metrological test items in the constancy test 31