

DIN ISO 9236-1:2007-08 (E)

Photography - Sensitometry of screen/film systems for medical radiography - Part 1: Determination of sensitometric curve shape, speed and average gradient (ISO 9236-1:2004)

Contents		Page
National foreword		4
National Annex NA (informative) Bibliography		5
Introduction		6
1	Scope	7
2	Normative references	7
3	Terms and definitions	7
4	General requirements	8
4.1	Storage and handling conditions	8
4.2	Safelights	8
4.3	X-ray equipment	9
4.4	Air kerma meter	9
4.5	Processing	9
4.6	Densitometry	10
5	Determination of sensitometric curve shape	10
5.1	General	10
5.2	Beam qualities	11
5.3	Geometry for curve shape determination	11
5.4	Exposure	14
5.5	Evaluation	14
6	Determination of average gradient	14
7	Determination of speed	16
7.1	Definition	16
7.2	Beam qualities	16
7.2.1	General	16
7.2.2	Phantom for Technique I	16
7.2.3	Phantom for Technique II	16
7.2.4	Phantom for Technique III	17
7.2.5	Phantom for Technique IV	17
7.2.6	Establishment of X-ray tube voltage	17
7.3	Geometry	19
7.4	Exposure	19
7.5	Evaluation	19
8	Speed and average gradient determination without sensitometric curve	22
9	Uncertainty	22
10	Test report	23
Annex A (informative) Rationale		24

A.1	Sensitometric curve shape	24
A.2	Speed	24
A.3	Phantoms	24
A.4	Air kerma	25
A.5	Dependence of speed on beam quality	25
Bibliography		26
Figures	Figure 1 -- Geometric set-up of the inverse-square-law sensitometer for exposure of the screen/film combination	12
	Figure 2 -- Geometric set-up for calibration of the inverse-square-law sensitometer	13
	Figure 3 -- Sensitometric curve	15
	Figure 4 -- Phantom for Technique I	17
	Figure 5 -- Geometric set-up for adjustment of beam qualities used for techniques I, II, III and IV	18
	Figure 6 -- Geometric set-up for the measurement of speed (Techniques I, II and III)	20
	Figure 7 -- Geometric set-up for the measurement of speed (Technique IV)	21
Tables	Table 1 -- Beam qualities for the determination of the sensitometric curve shape	11
	Table 2 -- Specification of techniques	16