

ISO 8185:2007-07 (E)

Respiratory tract humidifiers for medical use - Particular requirements for respiratory humidification systems

Contents	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	2
3 Terms and definitions	2
4 General requirements and general requirements for tests	4
5 Classification	4
6 Identification, marking and documents	5
7 Power input	7
8 Basic safety categories	7
9 Removable protective means	7
10 Environmental conditions	7
11 Not used	8
12 Not used	8
13 General	8
14 Requirements related to classification	8
15 Limitation of voltage and/or energy	8
16 Enclosures and protective covers	8
17 Separation	8
18 Protective earthing, functional earthing and potential equalization	8
19 Continuous leakage currents and patient auxiliary currents	8
20 Dielectric strength	9
21 Mechanical strength	9
22 Moving parts	9
23 Surface, corners and edges	9
24 Stability in normal use	9

25	Expelled parts	9
26	Vibration and noise	9
27	Pneumatic and hydraulic power	9
28	Suspended masses	9
29	X-Radiation	10
30	Alpha, beta, gamma, neutron radiation and other particle radiation	10
31	Microwave radiation	10
32	Light radiation (including lasers)	10
33	Infra-red radiation	10
34	Ultraviolet energy	10
35	Acoustical energy (including ultrasonics)	10
36	Electromagnetic compatibility	11
37	Locations and basic requirements	11
38	Marking, accompanying documents	11
39	Common requirements for category AP and category APG equipment	11
40	Requirements and tests for category AP equipment, parts and components thereof	11
41	Requirements and tests for category APG equipment, parts and components thereof	11
42	Excessive temperatures	11
43	* Fire prevention	12
44	Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization and disinfection	12
45	Pressure vessels and parts subject to pressure	13
46	Human error	13
47	Electrostatic charges	13
48	Biocompatibility	13
49	Interruption of power supply	13
50	Accuracy of operating data	14
51	* Protection against hazardous output	14
52	Abnormal operation and fault conditions	15
53	Environmental tests	16
54	General	16

55	Enclosure and covers	16
56	Components and general assembly	16
57	Mains parts, components and layout	17
58	Protective earthing -- Terminals and connections	17
59	Construction and layout	18
101	* Humidification system output	18
102	Liquid container	18
103	Alarm systems	18
Annex AA (informative) Rationale		20
Annex BB (normative) * Determination of the accuracy of the displayed temperature		27
Annex CC (informative) Specific enthalpy calculations		29
Annex DD (normative) Temperature sensors and mating ports		35
Annex EE (normative) * Determination of humidification system output		36
Annex FF (normative) * Standard temperature sensor		39
Annex GG (informative) Environmental aspects		41
Annex HH (informative) Reference to the essential principals of safety and performance		44
Annex II (informative) Terminology -- Index of defined terms		46
Bibliography		48