

ISO 22125-2:2019-11 (E)

Water quality - Technetium-99 - Part 2: Test method using inductively coupled plasma mass spectrometry (ICP-MS)

| Contents | | Page |
|--|--|-------------|
| Foreword | | iv |
| Introduction | | v |
| 1 | Scope | 1 |
| 2 | Normative references | 1 |
| 3 | Terms, definitions and symbols | 2 |
| 4 | Principle | 4 |
| 5 | Sampling, handling and storage | 4 |
| 6 | Procedure | 4 |
| 6.1 | Sample preparation for measurement | 4 |
| 6.2 | Sample measurement | 5 |
| 7 | Quality assurance and quality control program | 5 |
| 7.1 | General | 5 |
| 7.2 | Variables that could influence the measurement | 5 |
| 7.3 | Instrument verification | 5 |
| 7.4 | Contamination | 5 |
| 7.5 | Interference control | 6 |
| 7.6 | Method verification | 6 |
| 7.7 | Demonstration of analyst capability | 6 |
| 8 | Expression of results | 6 |
| 8.1 | Using Re, ⁹⁷ Tc, or ⁹⁸ Tc as a recovery tracer | 6 |
| 8.1.1 | Calculation of mass of tracer and analyte added | 6 |
| 8.1.2 | Measurement bias | 7 |
| 8.1.3 | Sample mass concentration | 7 |
| 8.1.4 | Detection limit | 8 |
| 8.1.5 | Limit of quantification | 8 |
| 8.2 | Using ^{95m} Tc, ^{97m} Tc or ^{99m} Tc as a recovery tracer | 8 |
| 8.2.1 | Calculation of activity of tracer, mass of analyte and mass of internal standard added | 8 |
| 8.2.2 | Purification step recovery | 9 |
| 8.2.3 | Measurement bias | 9 |
| 8.2.4 | Sample mass concentration | 9 |
| 8.2.5 | Detection limit | 9 |
| 8.2.6 | Limit of quantification | 10 |
| 8.2.7 | Conversion of mass concentration to activity concentration | 10 |
| 8.2.8 | Conversion of mass concentration to volume unit | 10 |
| 8.3 | Correction for the presence of ⁹⁹ Tc in the tracer | 11 |
| 9 | Test report | 11 |
| Annex A (informative) Method 1 -- TEVA resin | | 12 |
| Annex B (informative) Method 2 -- TRU resin | | 15 |

| | |
|---|-----------|
| Annex C (informative) Method 3 -- Anion exchange resin | 18 |
| Bibliography | 21 |