

DIN EN ISO 407:2023-12 (E)

Small medical gas cylinders - Pin-index yoke-type valve connections (ISO 40 7:2023)

Contents		Page
European foreword		3
Foreword		4
1	Scope	5
2	Normative references	5
3	Terms and definitions	5
4	Valve	5
5	Yoke	6
6	Basic dimensions	6
6.1	General.....	6
6.2	Pin-index yoke-type valve body.....	7
6.3	Single-pin yoke-type valve connection system.....	8
6.4	Two-pin yoke-type valve connection system with the pins in a single row.....	8
6.5	Two-pin yoke-type valve connection system with the pins in a double row.....	9
6.6	Basic dimensions.....	9
7	Requirements for alternative designs of yoke-type valve connections	10
7.1	Requirements for the design of the connecting yoke.....	10
7.2	Examples of alternative construction for the connecting yoke.....	10
8	Dimensions and positions of the holes and pins for yoke-type valve connections	13
8.1	General.....	13
8.2	Outlet connection with single-pin system.....	13
8.3	Outlet connections with two-pin/single-row system.....	14
8.3.1	Outlet connection for oxygen.....	14
8.3.2	Outlet connection for oxygen/carbon dioxide mixture ($\text{CO}_2 \leq 7\%$).....	15
8.3.3	Outlet connection for oxygen/helium mixture ($\text{He} \leq 80\%$).....	15
8.3.4	Outlet connection for ethylene.....	16
8.3.5	Outlet connection for nitrous oxide (with or without liquid draw-off).....	16
8.3.6	Outlet connection for cyclopropane.....	17
8.3.7	Outlet connection for helium and for helium/oxygen mixture ($\text{O}_2 < 20\%$).....	17
8.3.8	Outlet connection for carbon dioxide (with or without liquid draw-off) and for carbon dioxide/oxygen mixture ($\text{CO}_2 > 7\%$).....	18
8.3.9	Outlet connection for medical air.....	18
8.3.10	Outlet connection for nitrogen.....	19
8.4	Outlet connection with two-pin/double-row system.....	19