

ISO 16831:2012-04 (E)

Non-destructive testing - Ultrasonic testing - Characterization and verification of ultrasonic thickness measuring equipment

Contents		Page
Introduction		vi
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	General requirements for compliance	1
5	Manufacturer's technical specification for ultrasonic thickness measuring equipment	2
5.1	General	2
5.2	General attributes	2
5.3	Display	2
5.4	Transmitter	3
5.5	Receiver	3
5.6	Other information	3
6	Calibration blocks	3
6.1	General	3
6.2	Material	3
6.3	Shape and size	4
7	Performance requirements for ultrasonic thickness measuring equipment	5
8	Probes	6
9	Group 1 tests	7
9.1	General	7
9.2	Equipment required for Group 1 tests	7
9.3	Stability against temperature	7
9.4	Low battery warning	7
9.5	Battery operational time	8
9.6	Operational voltage range	8
9.7	Operational current range	8
9.8	Operational temperature range	9
9.9	Pulse repetition frequency, PRF	9
9.10	Transmitter pulse shape, rise-time and peak voltage	9
9.11	Receiver frequency range of operation	11
9.12	Minimum and maximum measurable thicknesses	11
9.13	Accuracy and resolution	11
9.14	Range of velocity setting (calibration)	11
9.15	Calibration mechanisms	11
9.16	Calibration setting storage	12
9.17	Data storage	12
9.18	Printing	12
9.19	Display and recall	13
9.20	Display response time	13
10	Group 2 tests	13
10.1	General	13

10.2	Equipment required for Group 2 tests	13
10.3	General characteristics	14
10.4	General mechanical state and external aspects	14
11	Group 3 tests	14
11.1	General	14
11.2	General mechanical state and external aspects	14
11.3	Calibration mechanisms	14
Foreword	v
11.4	Calibration setting storage	14
11.5	Data storage	15
Bibliography	16