

DIN 19266:2015-05 (E)

pH measurement - Reference buffer solutions for the calibration of pH measuring equipment

Contents	Page
Foreword	3
1 Scope	4
2 Normative references	4
3 Terms and definitions	4
4 Fundamentals	5
4.1 General	5
4.2 pH values of primary reference buffer solutions	5
4.3 pH values of secondary reference buffer solutions	6
5 Preparation of primary and secondary reference buffer solutions	9
5.1 General	9
5.2 Reference buffer solution A, pH(S) = 1,679 (25 °C), Potassium tetraoxalate	9
5.3 Reference buffer solution B, pH(S) = 3,557 (25 °C), Potassium hydrogen tartrate	9
5.4 Reference buffer solution C, pH(S) = 4,005 (25 °C), Potassium hydrogen phthalate	9
5.5 Reference buffer solution D, pH(S) = 6,865 (25 °C), Phosphate	9
5.6 Reference buffer solution E, pH(S) = 7,413 (25 °C), Phosphate	9
5.7 Reference buffer solution F, pH(S) = 9,180 (25 °C), Borax	9
5.8 Reference buffer solution G, pH(S) = 12,454 (25 °C), Calcium hydroxide	10
5.9 Reference buffer solution H, pH(S) = 3,776 (25 °C), Potassium dihydrogen citrate	10
5.10 Reference buffer solution I, pH(S) = 10,012 (25 °C), Sodium carbonate/sodium hydrogen carbonate	10
6 Storage and shelf life	10
7 Examples of pH(S) values of reference buffer solutions	10
8 Additional properties of reference buffer solutions	10
Annex A (informative) Examples of pH(S) values as a function of temperature at temperatures from 55 °C to 95 °C	12
Annex B (informative) Volumetric method	13
Bibliography	15