

# ISO 8299:2019 (E)

## Nuclear fuel technology — Determination of the isotopic and elemental uranium and plutonium concentrations of nuclear materials in nitric acid solutions by thermal-ionization mass spectrometry

---

### Contents

	Foreword
1	Scope
2	Normative references
3	Terms and definitions
4	Principle
5	Reference materials and reagents
5.1	Spikes and reference materials
5.2	Other chemical reagents
5.3	Resin, applicable for separation/purification of Pu and U
5.3.1	General
5.3.2	Preparation of resin
5.3.2.1	Anion exchange resin1 1 Anion exchange resins AG1 or AG2 are examples of suitable products supplied by Bio-rad. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the products named. Equivalent products may be used if they can be shown to lead to the same results.
5.3.2.2	Extraction separation resins2 2 TEVA from Eichrom or Tristen are examples of suitable products available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the products named. Equivalent products may be used if they can be shown to lead to the same results.
6	Apparatus
7	Apparatus for mass spectrometry
8	Sample preparation
8.1	Subsampling and spiking
8.1.1	Pellet or powder samples
8.1.2	Concentrated nuclear fuel solution samples (such as reprocessing solution)
8.1.3	Plutonium nitrate solution samples (such as product solution from a reprocessing plant)
8.1.4	Dried nitrate samples
8.2	Chemical valency adjustment
8.2.1	Valence adjustment with ferrous solution
8.2.2	Valence adjustment with hydrogen peroxide
8.3	Sample separation/purification
8.3.1	Ion exchange with anion-exchange resin
8.3.2	Purification with extraction separation resins (see 5.3.1.2)
8.4	Replicate treatments
9	Filaments preparation
9.1	Degassing of filaments
9.2	Sample loading
9.2.1	Normal sample loading
9.2.2	Graphite loading technique
9.2.3	Resin-bead loading on single filaments for Pu samples
9.3	Filament mounting (filament assemblies preparation)

- 10 Instrument calibration
    - 10.1 Mass calibration
    - 10.2 Gain calibration for Faraday multi-detectors
    - 10.3 Faraday detector calibration
    - 10.4 Mass discrimination calibration
  - 11 Isotopic mass spectrometric measurements
    - 11.1 Total evaporation measurements using a single or double filament assembly and a multi-Faraday collector system
    - 11.2 Bias correction method using a double filament assembly and a multi-Faraday collector system
  - 12 Calculation of the results
    - 12.1 Calculation of ion current intensities
    - 12.2 Calculation of mean, weighted mean and standard deviation on a set of ratios  $x_i$ , ( $i = 1 \dots N$ )
    - 12.3 Mass discrimination correction
    - 12.4 Calculation of the atomic percent abundance  $A_i$
    - 12.5 Calculation of the isotopic mass percent  $W_j$
    - 12.6 Calculation of concentration
    - 12.7 Isotope decay correction
  - 13 Blanks
  - 14 Quality control
  - 15 Measurement uncertainty
    - 15.1 Elemental assay
    - 15.2 Isotopic analysis
  - 16 Interferences
- Annex A (normative) Preparation and standardization of spike solutions**
- A.1 General
  - A.2 Principle
  - A.3 Standard solutions of certified chemical reference materials
    - A.3.1 Stock solution of uranium standard reference solution
    - A.3.2 Stock solution of plutonium standard reference solution
  - A.4 Preparation of spike solutions
    - A.4.1 Solution of U spike
    - A.4.2 Solution of Pu spike
    - A.4.3 Mixed solution of US and PuS spikes
    - A.4.4 Preparation of LSD spikes
  - A.5 Standardization of the dilute spike solutions by isotope dilution mass spectroscopy
    - A.5.1 Standardization of the uranium spike
      - A.5.1.1 Isotopic analysis of the diluted solutions of certified chemical standard
      - A.5.1.2 Isotopic analysis of the spike solution
      - A.5.1.3 Standardization of the US isotope concentration in the spike solution
    - A.5.2 Standardization of the plutonium spike
      - A.5.2.1 Isotopic analysis of the diluted solutions of certified chemical standard
      - A.5.2.2 Isotopic analysis of the spike solution
      - A.5.2.3 Standardization of the PuS isotope concentration in the spike solution
    - A.5.3 Standardization of mixed spike solutions
    - A.5.4 Standardization of isotope dilution analysis with joint spikes
  - A.6 Validation of LSD spikes
    - A.6.1 Validation of the isotopic composition
    - A.6.2 Validation of the element concentrations