ISO 23449:2020 (E)

Corrosion of metals and alloys — Multielectrode arrays for corrosion measurement

Contents

Foreword

Introduction

1 Scope

2 Normative references

3 Terms and definitions

4 Principle

- 4.1 Multielectrode arrays
- 4.2 Coupled multielectrode array (CMA)
- 4.3 Multielectrode array with closely packed electrodes for studying spatiotemporal behaviour of localized corrosion
- 4.4 Coupled multielectrode array sensor (CMAS)
- 4.4.1 CMAS for corrosion monitoring
- 4.4.2 CMAS used without polarization to measure corrosion rate at free corrosion potential
- 4.4.3 CMAS used to evaluate the effectiveness of cathodic protection and the effect of stray current
- 4.5 Multielectrode arrays for high throughput measurements
- 4.6 Multielectrode arrays for other applications

5 Instrumentation

- 5.1 Potential measurement
- 5.2 Coupling current measurement
- 5.3 Effective coupling of individual electrodes
- 5.3.1 Coupling with multichannel ZVA
- 5.3.2 Coupling with wires and measuring current with a single ZVA

6 Fabrication of multielectrode array

- 6.1 Electrode preparation
- 6.2 Number of electrodes
- 6.3 Mounting of electrodes
- 6.4 Surface coating on electrodes for preventing crevice corrosion
- 6.5 Electrode configuration
- 6.6 Size of electrodes
- 6.7 Spacing of electrodes for spatiotemporal studies
- 6.8 Spacing of electrodes for corrosion monitoring in oil and gas application
- 6.9 Size and spacing of the electrodes for high throughput studies

7 Test procedure

8 Test report

Annex A (informative) Typical results from multielectrode array with closely packed electrodes for studying spatiotemporal behaviour of localized corrosion

Annex B (informative) Typical results from a CMAS for corrosion monitoring

- B.1 Typical response of the CMAS to different environments
- B.2 Typical results from CMAS under cathodic protection conditions

Annex C (informative) Example reports

Page count: 20