

ISO 23468:2021 (E)

Reactor technology — Power reactor analyses and measurements — Determination of heavy water isotopic purity by Fourier transform infrared spectroscopy

Contents

	Foreword
1	Scope
2	Normative references
3	Terms and definitions
4	Principle
5	Reagents and materials
6	Apparatus
7	Heavy water reference standard preparation
7.1	General requirements
7.2	Storage of high isotopic purity heavy water reference standard
7.2.1	Ampoule bottle
7.2.2	Specially designed stainless steel container
7.3	Lower D ₂ O concentration reference standard preparation
8	Sampling
8.1	General requirements
8.2	Sampling with syringes
8.3	Sampling with bottles
9	Preparation of calibration curves of the method
9.1	General
9.2	Preparation of the apparatus
9.2.1	FTIR spectrometer
9.2.2	Infrared liquid cell
9.3	Editing of the method
9.4	Background spectrum
9.5	Heavy water reference standard spectra
9.6	Draw the calibration curves for different D ₂ O concentration ranges
10	Procedure
10.1	Sample preparation
10.1.1	Sample temperature adjustment
10.1.2	Sample filtration
10.1.3	Sample source check
10.2	Sample load and scan
10.3	After measurement
11	Expression of results
11.1	Calculation method
11.2	Precision
11.3	Uncertainty
12	Interferences
12.1	Contaminated samples
12.2	Air bubbles

12.3 Foreign materials on the outside of the cell window

13 Test report

Annex A (informative) Specially designed stainless steel container for D2O reference standard

Annex B (informative) Interested region suggested for typical curves

Page count: 19