ISO 24417:2022-09 (E)

Surface chemical analysis - Analysis of metallic nanolayers on iron based substrates by glow-discharge optical-emission spectrometry

Contents

Page

Foreword ..... **v**

1 Scope **1**

2 Normative references **1**

3 Terms and definitions **1**

4 Principle **1**

5 Apparatus **2**
5.1 Glow discharge optical emission spectrometer 2
5.1.1 General 2
5.1.2 Selection of spectral lines 2
5.1.3 Selection of glow discharge source type 2

6 Adjusting the glow discharge spectrometer system settings **3**
6.1 General 3
6.2 Setting the parameters of a DC source 4
6.2.1 Constant applied current and voltage 4
6.2.2 Constant applied current and pressure 5
6.2.3 Constant voltage and pressure 5
6.3 Setting the discharge parameters of an RF source 6
6.3.1 General 6
6.3.2 Constant applied power and pressure 6
6.3.3 Constant applied power and DC bias voltage 6
6.3.4 Constant effective power and effective RF voltage 7
6.4 Minimum performance requirements 7
6.4.1 General 7
6.4.2 Control of lamp cleanliness and start-up performance 7
6.4.3 Minimum repeatability 9
6.4.4 Detection limit 9

7 Sampling **11**

8 Calibration **11**
8.1 General 11
8.2 Calibration samples 11
8.2.1 General 11
8.2.2 Low alloy iron or steel samples 12
8.2.3 Stainless steel samples 12
8.2.4 Nickel alloy samples 12
8.2.5 Titanium alloy samples 12
8.2.6 Aluminium alloy samples 12
8.2.7 High oxygen samples 12
8.2.8 High carbon samples 12
8.2.9 High purity copper sample 12
8.3 Validation samples 13
8.4 Determination of the sputtering rate of calibration samples 13
8.5 Emission intensity measurements of calibration samples 14
8.6 Calculation of calibration equations 14
8.7 Validation using reference materials 14
8.7.1 General 14
8.7.2 Checking analytical accuracy using bulk reference materials 15
8.7.3 Checking analytical accuracy using surface layer reference materials 15

8.8 Verification and drift correction 15
9 Analysis of test specimens **16**
9.1 Adjusting discharge parameters 16
9.2 Setting of measuring time and data acquisition rate 16
9.3 Quantifying depth profiles of test specimens 16
10 Expression of results **17**
10.1 Expression of quantitative depth profile 17
10.2 Determination of total coating mass per unit area (coating areic mass) 17
11 Precision **17**
12 Test report **18**
Annex A (normative) **Calculation of calibration constants and quantitative evaluation of depth profiles** **19**
Annex B (informative) **Suggestions concerning suitable spectral lines** **31**
Annex C (informative) **Additional information on international cooperative tests** **32**
Bibliography **35**