ISO 4266-5:2024-09 (E)

Petroleum and liquid petroleum products - Measurement of level and temperature in storage tanks by automatic methods - Part 5: Measurement of temperature in marine vessels

Contents

Page

Foreword iv

1 Scope 1

2 Normative reference 1

3 Termsanddefinitions 1

4 Precautions 2
4.1 Safety precautions 2
4.2 Equipment precautions 2
4.3 General precautions 2

5 Accuracy 3
5.1 General 3
5.2 Intrinsic error of ATTs 3
5.3 Calibration prior to installation 3
5.3.1 General 3
5.3.2 ATT calibrated as a system 4
5.3.3 ATT calibrated by components 4
5.3.4 Multiple-point ATTs 4
5.3.5 Uncertainty of the reference 4
5.4 Error caused by installation and operating conditions 4
5.5 Overall accuracy 4
5.5.1 General 4
5.5.2 Use of ATT for fiscal/custody transfer purposes 5

6 Selection of ATTs 5
6.1 General 5
6.2 ATTs for fiscal/custody transfer purpose 5

7 Description of ATT equipment -- Electrical temperature elements 6
7.1 General 6

8 Installation of marine ATTs 6
8.1 General 6
8.2 Location of temperature-sensing element(s) 6
8.3 Location of manual (temperature) gauge hatch 6
8.4 Temperature measurement of inerted tanks 6

9 CalibrationandOn-boardverificationofmarineATTs 7
9.1 General 7
9.2 Calibration of single-point temperature element ATTs for fiscal/custody transfer purposes 7
9.2.1 Calibration prior to installation 7
9.2.2 Initial verification at shipyard or during sea trials 7
9.3 Calibration of multiple-point ATTs for fiscal/custody transfer purposes 8
9.3.1 Calibration prior to installation 8
9.3.2 Initial on-board verification at shipyard or during sea trials 8

9.4 Subsequent verification of ATTs 9
9.4.1 General 9
9.4.2 Frequency of subsequent verification 9
9.4.3 Record keeping 9

10 Data communication and receiving 10

Annex A (informative) Accuracy limitations of tank-temperature measurements on board marine vessels 11

Bibliography 12