

DIN EN ISO 8536-4:2011-01 (E)

Infusion equipment for medical use - Part 4: Infusion sets for single use, gravity feed (ISO 8536-4:2010)

Contents	Page
Foreword	3
1 Scope	4
2 Normative references	4
3 General requirements	4
4 Designation	7
4.1 Infusion set	7
4.2 Air-inlet device	7
5 Materials	7
6 Physical requirements	8
6.1 Particulate contamination	8
6.2 Leakage	8
6.3 Tensile strength	8
6.4 Closure-piercing device	8
6.5 Air-inlet device	8
6.6 Tubing	9
6.7 Fluid filter	9
6.8 Drip chamber and drip tube	9
6.9 Flow regulator	9
6.10 Flow rate of infusion fluid	9
6.11 Injection site	9
6.12 Male conical fitting	9
6.13 Protective caps	9
7 Chemical requirements	10
7.1 Reducing (oxidizable) matter	10
7.2 Metal ions	10
7.3 Titration acidity or alkalinity	10
7.4 Residue on evaporation	10
7.5 UV absorption of extract solution	10
8 Biological requirements	10
8.1 General	10
8.2 Sterility	10
8.3 Pyrogenicity	10
8.4 Haemolysis	10
8.5 Toxicity	11
9 Labelling	11
9.1 Unit container	11
9.2 Shelf or multi-unit container	11
10 Packaging	12
Annex A (normative) Physical tests	13

Annex B (normative) Chemical tests	17
Annex C (normative) Biological tests	19
Bibliography	20
Annex ZA (informative) Relationship between this International Standard and the Essential Requirements of EU Directive 93/42/EEC on Medical devices	21