

DIN EN ISO 10079-1:2016-05 (E)

Medical suction equipment - Part 1: Electrically powered suction equipment (ISO 10079-1:2015), German version EN ISO 10079-1:2015

Contents	Page
European foreword	4
Annex ZA (informative) Relationship between this European Standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered	5
Foreword	8
1 Scope	9
2 Normative references	9
3 Terms and definitions	10
4 General requirements	12
4.1 Risk management	12
4.2 Usability	13
4.3 Clinical investigation	13
4.4 Biophysical or modelling research	13
4.5 Test methods	13
5 Cleaning, disinfection and sterilization	13
6 Design requirements	14
6.1 Collection container	14
6.1.1 General	14
6.1.2 Container capacity	14
6.1.3 Container strength	14
6.2 Connections	14
6.2.1 Tubing connectors for collection containers	14
6.2.2 Inlet port	15
6.2.3 Exhaust port	15
6.3 Suction tubing	15
6.4 Vacuum level indicators	15
6.5 Spillage on electrical suction equipment	16
7 Operational requirements	16
7.1 Ease of operation	16
7.2 Dismantling and reassembly	16
7.3 Mechanical shock	16
7.4 Stability	16
7.5 Protective devices	17
7.5.1 Contamination protection	17
7.5.2 Overfill protection devices	17
7.5.3 Pressure protection	17
7.6 Noise	17
7.6.1 Low vacuum/low flowrate equipment	17
7.6.2 Suction equipment other than that specified in 7.6.1	17
7.7 Air leakage	18
7.7.1 Collection containers for general use	18
7.7.2 Collection containers for thoracic drainage	18

8	Physical requirements for suction equipment for field use	18
8.1	(*) Dimensions	18
8.2	Mass	18
9	Performance requirements for vacuum level and flowrate	18
9.1	High vacuum/high flowrate equipment	18
9.2	Medium vacuum equipment	19
9.3	Low vacuum/low flowrate equipment	19
9.4	Low vacuum/high flowrate equipment	19
9.5	Thoracic drainage equipment for adults	19
9.6	Intermittent vacuum equipment	19
9.7	Vacuum regulators with fixed setting	19
9.8	Vacuum regulators with variable setting	20
9.9	Equipment intended for pharyngeal suction	20
9.10	Battery powered transportable suction equipment	20
9.11	Interruption of the power supply	20
10	(*) Resistance to environment of suction equipment for field and/or transport use	20
10.1	Operating conditions	20
10.2	Storage	20
11	Information to be supplied by the manufacturer (labelling and instructions for use)	21
11.1	Information supplied by the manufacturer shall comply with EN 1041	21
11.3	Labelling of equipment	21
11.4	Instructions for use	22
	Annex A (normative) Test methods	24
	Annex B (informative) Rationale statement	35
	Annex C (informative) Lumen size and its effect on flowrate	36
	Annex D (informative) Schematic of suction equipment	37
	Bibliography	38