

DIN EN ISO 1135-4:2016-06 (E)

Transfusion equipment for medical use - Part 4: Transfusion sets for single use, gravity feed (ISO 1135-4:2015)

Contents	Page
European Foreword	4
Foreword	6
1 Scope	7
2 Normative references	7
3 General requirements	8
3.1 Nomenclature for components of the transfusion set	8
3.2 Maintenance of sterility	8
4 Materials	9
5 Physical requirements	9
5.1 Particulate contamination	9
5.2 Leakage	9
5.3 Tensile strength	9
5.4 Closure-piercing device	9
5.5 Tubing	10
5.6 Filter for blood and blood components	10
5.7 Drip chamber and drip tube	10
5.8 Flow regulator	10
5.9 Flow rate of blood and blood components	10
5.10 Injection site	11
5.11 Male conical fitting	11
5.12 Protective caps	11
6 Chemical requirements	11
6.1 Reducing (oxidizable) matter	11
6.2 Metal ions	11
6.3 Titration acidity or alkalinity	11
6.4 Residue on evaporation	11
6.5 UV absorption of extract solution	11
7 Biological requirements	12
7.1 General	12
7.2 Sterility	12
7.3 Pyrogenicity	12
7.4 Haemolysis	12
7.5 Toxicity	12
7.6 Assessment of blood component depletion	12
7.7 Assessment of damage to blood components	12
8 Labelling	13
8.1 General	13
8.2 Unit container	13
8.3 Shelf or multi-unit container	13
9 Packaging	14

10	Disposal	14
	Annex A (normative) Physical tests	15
	Annex B (normative) Chemical tests	19
	Annex C (normative) Biological tests	21
	Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on Medical devices	22
	Bibliography	24