DIN EN 14683:2019-10 (E)

Medical face masks - Requirements and test methods (includes Amendment :2019)

Contents

Page

European foreword 4

Introduction 5

1 Scope 6

2 Normative references 6

3 Terms and definitions 6

4 Classification 8

5 Requirements 8
5.1 General 8
5.1.1 Materials and construction 8
5.1.2 Design 8
5.2 Performance requirements 8
5.2.1 General 8
5.2.2 Bacterial filtration efficiency (BFE) 8
5.2.3 Breathability 8
5.2.4 Splash resistance 8
5.2.5 Microbial cleanliness (Bioburden) 9
5.2.6 Biocompatibility 9
5.2.7 Summary of performance requirements 9

6 Marking, labelling and packaging 9

Annex A (informative) Information for users 11

Annex B (normative) Method for in vitro determination of bacterial filtration efficiency (BFE) 12

B.1 General 12
B.2 Principle 12
B.3 Reagents and materials 12
B.3.1 General 12
B.3.2 Tryptic soy agar 12
B.3.3 Tryptic soy broth 12
B.3.4 Peptone water 13
B.3.5 Culture of Staphylococcus aureus ATCC 6538, growing on tryptic soy agar slants 13
B.4 Test apparatus 13
B.4.1 Six stage cascade impactor, the arrangement is specified in Table B.1 13
B.4.2 Nebulizer, capable of delivering particles with a mean size of (3,0 ± 0,3) m when in contact with the cascade impactor 13
B.4.3 Aerosol chamber, glass, 600 mm long and 80 mm in external diameter 13
B.4.4 Flow meters, capable of measuring a flow rate of 28,3 l/min 13
B.4.5 Pressure gauge, capable of measuring a pressure of 35 kPa to an accuracy of ± 1 kPa 13
B.4.6 Erlenmeyer flasks, 250 ml and 500 ml capacity 13
B.4.7 Peristaltic or syringe pump, capable of delivering 0,01 ml/min 13
B.4.8 Vacuum pump, capable of maintaining a flow rate of 57 l/min 13
B.5 Test specimens 13
B.6 Preparation of bacterial challenge 13
B.7 Procedure 14

B.8 Calculation of bacterial filtration efficiency (BFE) 15
B.9 Test report 16

Annex C (normative) Method for determination of breathability (differential pressure) 18

C.1 Principle 18
C.2 Test apparatus 19
C.2.1 Mass flow meter(s) capable of measuring an airflow of 8 l/min 19
C.2.2 Manometer, a differential manometer (water or digital). Individual manometers can also be used. M1 is for the upstream pressure measurement and M2 is for the downstream pressure measurement 19
C.2.3 Electric vacuum pump including a pressure buffer tank 19
C.2.4 Valve permitting the adjustment of the flow rate 19
C.2.5 Sample holder 19
C.3 Test specimens 19
C.4 Procedure 20
C.5 Calculation of differential pressure 20
C.6 Test report 20

Annex D (informative) Microbial cleanliness 21

D.1 Sampling 21
D.2 Testing 21

Annex ZA (informative) Relationship between this European Standard and the essential requirements of Directive 93/42/EEC [1993 OJ L 169] aimed to be covered 22

Bibliography 23