

ISO 11140-3:2007-03 (E)

Sterilization of health care products - Chemical indicators - Part 3: Class 2 indicator systems for use in the Bowie and Dick-type steam penetration test

Contents	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 General requirements	2
5 Indicator system format	2
6 Performance requirements	2
7 Packaging and labelling	3
8 Quality assurance	4
Annex A (normative) Determination of strength after steam sterilization	5
Annex B (normative) Estimation of visual difference between colour of the substrate and the changed (or unchanged) indicator system by determination of relative reflectance density	6
Annex C (normative) Determination of indicator colour change on exposure to dry saturated steam	10
Annex D (normative) Determination of indicator colour change on exposure to dry heat	11
Annex E (normative) Accelerated ageing of test samples	12
Annex F (normative) Determination of transfer of indicator to standard test pack on processing	13
Annex G (normative) Determination of shelf life of the product	14
Annex H (normative) Steam exposure apparatus	15
Annex I (normative) Determination of sensitivity of the indicator to the presence of air	17
Annex J (normative) Air injection system	18
Annex K (normative) Standard test pack	20
Bibliography	21