

ISO 27186:2010-03 (E)

Active implantable medical devices - Four-pole connector system for implantable cardiac rhythm management devices - Dimensional and test requirements

| Contents | | Page |
|--|--|-------------|
| Foreword | | iv |
| Introduction | | v |
| 1 Scope | | 1 |
| 2 Normative references | | 1 |
| 3 Terms and definitions | | 2 |
| 4 Requirements | | 5 |
| Annex A (normative) Electrical isolation test | | 22 |
| Annex B (informative) Rationale for Annex A | | 27 |
| Annex C (normative) Dielectric strength test | | 29 |
| Annex D (informative) Rationale for Annex C | | 35 |
| Annex E (normative) Current-carrying test high voltage types | | 39 |
| Annex F (informative) Rationale for Annex E | | 44 |
| Annex G (informative) Lead connector fatigue strength test | | 46 |
| Annex H (informative) Lead connector seal zone materials | | 47 |
| Annex I (informative) Seal zone creep | | 49 |
| Annex J (informative) Contact resistance stability | | 54 |
| Annex K (informative) Rationale for Annex J | | 58 |
| Annex L (informative) Selection of contact materials | | 60 |
| Annex M (normative) Lead connector contact material requirements | | 62 |
| Annex N (informative) Rationale for Annex M | | 66 |
| Annex O (informative) Rationale for requirements in this International Standard | | 72 |
| Annex P (informative) Connector products (e.g. adaptors, extenders, patient cables, etc.) | | 79 |
| Bibliography | | 81 |