

ISO 14708-2:2012-08 (E)

Implants for surgery - Active implantable medical devices - Part 2: Cardiac pacemakers

Contents		Page
Foreword		v
Introduction		vi
1	Scope	1
2	Normative references	1
3	Terms and definitions	2
4	Symbols and abbreviated terms	6
5	General requirements for non-implantable parts	6
6	Measurements of implantable pulse generator and lead characteristics	6
6.1	Measurement of implantable pulse generator characteristics	6
6.2	Measurement of the lead pacing impedance (Z_p)	18
7	General arrangement of the packaging	19
8	General markings for active implantable medical devices	20
9	Markings on the sales packaging	20
10	Construction of the sales packaging	21
11	Markings on the sterile pack	21
12	Construction of the non-reusable pack	22
13	Markings on the active implantable medical device	23
14	Protection from unintentional biological effects being caused by the active implantable medical device	24
15	Protection from harm to the patient or user caused by external physical features of the active implantable medical device	24
16	Protection from harm to the patient caused by electricity	24
17	Protection from harm to the patient caused by heat	25
18	Protection from ionizing radiation released or emitted from the active implantable medical device	26
19	Protection from unintended effects caused by the device	26
20	Protection of the device from damage caused by external defibrillators	27
21	Protection of the device from changes caused by high power electrical fields applied directly to the patient	27

22	Protection of the active implantable medical device from changes caused by miscellaneous medical treatments	27
23	Protection of the active implantable medical device from mechanical forces	27
24	Protection of the active implantable medical device from damage caused by electrostatic discharge	32
25	Protection of the active implantable medical device from damage caused by atmospheric pressure changes	32
26	Protection of the active implantable medical device from damage caused by temperature changes	33
27	Protection of the active implantable medical device from electromagnetic non-ionizing radiation	33
28	Accompanying documentation	33
Annex A A (informative) Relationship between the fundamental principles in ISO/TR 14283 and the clauses of this International Standard		37
fundamental principles in Annex A		51
Annex C C (informative) Rationale		52
Annex D D (informative) Code for describing modes of implantable pulse generators		61
Annex E E (informative) Symbols		64
Annex F F (normative) Pulse forms		65
Bibliography		67