

ISO 80601-2-12:2020-02 (E)

Medical electrical equipment - Part 2-12: Particular requirements for basic safety and essential performance of critical care ventilators

Contents	Page
201.1 Scope, object and related standards	1
201.1.1 * Scope.....	1
201.1.2 Object.....	2
201.1.3 Collateral standards.....	3
201.1.4 Particular standards.....	3
201.2 Normative references	4
201.3 Terms and definitions	7
201.4 General requirements	9
201.4.3 <i>Essential performance</i>	9
201.4.3.101 * Additional requirements for <i>essential performance</i>	9
201.4.4 Additional requirements for <i>expected service life</i>	9
201.4.6 * <i>ME equipment</i> or <i>ME system</i> parts that contact the <i>patient</i>	10
201.4.11.101 * Additional requirements for pressurized gas input.....	10
201.4.11.101.1 Overpressure requirement.....	10
201.4.11.101.2 Compatibility requirement.....	10
201.5 General requirements for testing of <i>ME equipment</i>	11
201.5.101 Additional requirements for general requirements for testing of <i>ME equipment</i>	11
201.5.101.1 <i>Ventilator</i> test conditions.....	11
201.5.101.2 * Gas flowrate and leakage specifications.....	11
201.5.101.3 * <i>Ventilator</i> testing errors.....	11
201.6 Classification of <i>ME equipment</i> and <i>ME systems</i>	12
201.7 <i>ME equipment</i> identification, marking and documents	12
201.7.2.3 * Consult <i>accompanying documents</i>	12
201.7.2.4.101 Additional requirements for <i>accessories</i>	12
201.7.2.13.101 Additional requirements for physiological effects.....	12
201.7.2.17.101 Additional requirements for protective packaging.....	12
201.7.2.18 External gas source.....	13
201.7.2.101 * Additional requirements for marking on the outside of <i>ME equipment</i> or <i>ME equipment</i> parts.....	13
201.7.4.3 * Units of measurement.....	14
201.7.9.1 Additional general requirements.....	14
201.7.9.2.1.101 Additional general requirements.....	14
201.7.9.2.2.101 * Additional requirements for warnings and safety notices.....	15
201.7.9.2.8.101 * Additional requirements for start-up <i>procedure</i>	16
201.7.9.2.9.101 * Additional requirements for operating instructions.....	16
201.7.9.2.12 <i>Cleaning, disinfection, and sterilization</i>	17
201.7.9.2.14.101 * Additional requirements for <i>accessories</i> , supplementary equipment, used material.....	17

201. 7.9.2.16.101	* Additional requirements for reference to the technical description	18
201. 7.9.3.1.101	* Additional general requirements	18
201. 7.9.3.101	Additional requirements for the technical description	18
201. 8	Protection against electrical <i>hazards</i> from <i>ME equipment</i>.....	19
201. 9	Protection against <i>mechanical hazards</i> of <i>ME equipment</i> and <i>ME systems</i>	19
201. 9.6.2.1.101	* Additional requirements for audible acoustic energy.....	19
201. 9.101	* Additional requirements for suction <i>procedures</i>	20
201. 10	Protection against unwanted and excessive radiation <i>hazards</i>.....	23
201. 11	Protection against excessive temperatures and other <i>hazards</i>	23
201. 11.1.2.2	* <i>Applied parts</i> not intended to supply heat to a <i>patient</i>	23
201. 11.6.5.101	* Additional requirements for ingress of water or particulate matter into <i>ME equipment</i> or <i>ME system</i>	23
201. 11.6.6	* <i>Cleaning and disinfection</i> of <i>ME equipment</i> or <i>ME system</i>	24
201. 11.6.7	<i>Sterilization</i> of <i>ME equipment</i> or <i>ME system</i>	24
201. 11.7	<i>Biocompatibility</i> of <i>ME equipment</i> and <i>ME systems</i>	24
201. 11.8.101	* Additional requirements for interruption of the power supply/ <i>supply mains</i> to <i>ME equipment</i>	25
201. 12	Accuracy of controls and instruments and protection against hazardous outputs	27
201. 12.1	* Accuracy of controls and instruments	27
201. 12.1.101	* <i>Volume-control inflation-type</i>	27
201. 12.1.102	* <i>Pressure-control inflation-type</i>	31
201. 12.1.103	Other <i>inflation-types</i>	34
201. 12.1.104	* <i>Inspiratory volume monitoring</i>	35
201. 12.1.105	* Response of the <i>ventilator</i> to an increase in set O ₂ concentration	35
201. 12.4	Protection against hazardous output	37
201. 12.4.101	Oxygen monitor	37
201. 12.4.102	* Measurement of <i>airway pressure</i>	38
201. 12.4.103	* Measurement of expired volume and low volume <i>alarm conditions</i>	39
201. 12.4.103.1	<i>Ventilators</i> intended to provide a <i>tidal volume</i> >50 ml.....	39
201. 12.4.103.2	<i>Ventilators</i> intended to provide a <i>tidal volume</i> ≤50 ml.....	40
201. 12.4.104	* Expiratory end-tidal CO ₂ <i>monitoring equipment</i>	41
201. 12.4.105	* <i>Maximum limited pressure protection device</i>	42
201. 12.4.106	* High <i>airway pressure alarm condition</i> and <i>protection device</i>	42
201. 12.4.107	<i>PEEP alarm conditions</i>	43
201. 12.4.108	* <i>Obstruction alarm condition</i>	44
201. 12.4.109	* <i>Disconnection alarm condition</i>	45
201. 12.4.110	Protection against inadvertent setting of high <i>airway pressure</i>	45
201. 12.101	* Protection against accidental or unintentional adjustments	45
201. 13	<i>Hazardous situations and fault conditions for ME equipment</i>	46

201. 13.2.101	* Additional specific <i>single fault conditions</i>	46
201. 13.2.102	* Failure of one gas supply to a <i>ventilator</i>	46
201. 13.2.103	* Independence of ventilation control function and related <i>risk control</i> measures.....	47
201. 13.2.104	* Failure of <i>functional connection</i> to a <i>ventilator</i> control or monitoring means.....	47
201. 14	<i>Programmable electrical medical systems (PEMS)</i>	47
201. 14.101	Software life cycle	48
201. 15	Construction of <i>ME</i> equipment	48
201. 15.3.5.101	Additional requirements for rough handling.....	48
201. 15.3.5.101.1	* Shock and vibration (robustness)	48
201. 15.3.5.101.2	* Shock and vibration for a <i>transit-operable ventilator</i> during operation	49
201. 15.4.1	Construction of connectors	51
201. 15.101	Mode of operation	51
201. 15.102	Delivered oxygen concentration.....	51
201. 15.103	<i>Accessory</i> self-check.....	51
201. 16	<i>ME</i> systems	52
201. 16.1.101	Additional general requirements for <i>ME</i> systems	52
201. 16.2.101	* Additional general requirements for <i>accompanying documents</i> of an <i>ME</i> system	52
201. 17	Electromagnetic compatibility of <i>ME</i> equipment and <i>ME</i> systems	52
201. 101	Gas connections	52
201. 101.1	* Protection against reverse gas leakage	52
201. 101.2	Connection to a <i>high-pressure input port</i>	53
201. 101.2.1	Connector	53
201. 101.2.2	* Filter.....	53
201. 101.3	<i>VBS</i> connectors.....	53
201. 101.3.1	* General	53
201. 101.3.2	Other named ports	53
201. 101.3.2.1	<i>Patient-connection port</i>	53
201. 101.3.2.2	<i>Gas output port</i> and <i>gas return port</i>	54
201. 101.3.2.3	<i>Emergency intake port</i>	54
201. 101.3.2.4	<i>Flow-direction-sensitive components</i>	54
201. 101.3.2.5	* <i>Accessory</i> port.....	54
201. 101.3.2.6	<i>Gas exhaust port</i>	55
201. 101.3.2.7	Temperature sensor port.....	55
201. 102	Requirements for the <i>VBS</i> and <i>accessories</i>	55
201. 102.1	* General	55
201. 102.2	Labelling	55
201. 102.3	Breathing tubes.....	55
201. 102.4	* Water vapour management.....	56
201. 102.4.1	Humidification system	56
201. 102.4.2	<i>Heat and moisture exchanger (HME)</i>	56

201. 102.6	<i>Breathing system filters</i>	56
201. 102.7	<i>Ventilator breathing systems</i>	56
201. 102.7.1	* Leakage from complete <i>VBS</i>	56
201. 102.7.2	* Non-invasive ventilation	57
201. 103	* Spontaneous breathing during loss of power supply	57
201. 104	* Indication of duration of operation	57
201. 105	<i>Functional connection</i>	58
201. 105.1	General	58
201. 105.2	* Connection to an electronic health record	58
201. 105.3	* Connection to a <i>distributed alarm system</i>	58
201. 105.4	Connection for remote control	58
201. 106	Display loops	58
201. 106.1	Pressure-volume loops	58
201. 106.2	Flow-volume loops	59
201. 107	* Timed ventilatory pause	59
201. 107.1	<i>Expiratory pause</i>	59
201. 107.2	<i>Inspiratory pause</i>	60
202	Electromagnetic disturbances — Requirements and tests	61
206	Usability	62
206.101	<i>Primary operating functions</i>	62
206.102	* Training	63
208	General requirements, tests and guidance for <i>alarm systems in medical electrical equipment and medical electrical systems</i>	64
Annex C (informative)	Guide to marking and labelling requirements for <i>ME equipment and ME systems</i>	66
201.C.101	Marking on the outside of <i>ME equipment, ME systems</i> or their parts	66
201.C.102	<i>Accompanying documents</i> , general	67
201.C.103	<i>Accompanying documents</i> , instructions for use	67
201.C.104	<i>Accompanying documents</i> , technical description	70
Annex D (informative)	Symbols on marking	71
Annex AA (informative)	Particular guidance and rationale	75
AA.1	General guidance	75
AA.2	Rationale for particular clauses and subclauses	75
Annex BB (informative)	Data interfaces	114
BB.1	Background and purpose	114
BB.2	Data definition	115
Annex CC (informative)	Reference to the essential principles	123
Annex DD (informative)	Reference to the general safety and performance requirements	126
Annex EE (informative)	Terminology — Alphabetized index of defined terms	129
Bibliography	134