IEC 80601-2-52:2026-05 (E)

Medical electrical equipment - Part 2-52: Particular requirements for the basic safety and essential performance of medical beds for adults

Contents

Page

FOREWORD 4
INTRODUCTION 7
201.1 Scope, object and related standards 8
201.2 Normative references 10
201.3 Terms and definitions 10
201.4 General requirements 15
201.5 General requirements for testing of ME EQUIPMENT 16
201.6 Classification of ME EQUIPMENT and ME SYSTEMS 19
201.7 ME EQUIPMENT identification, marking and documents 19
201.8 Protection against electrical HAZARDS from ME EQUIPMENT 25
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS 26
201.10 Protection against unwanted and excessive radiation HAZARDS 53
201.11 Protection against excessive temperatures and other HAZARDS 53
201.12 Accuracy of controls and instruments and protection against hazardous outputs 57
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT 57
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) 58
201.15 Construction of ME EQUIPMENT 58
201.16 ME SYSTEMS 63
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS 63
Annexes 64
Annex AA (informative) Particular guidance and rationale 65
Annex BB (normative) Additional design requirements and recommendations for MEDICAL BEDS 84
Annex CC (informative) Particular guidance for assessing RISK of entrapment in V-shaped openings 89
Annex DD (informative) Guidance and recommendations for periodic inspection 95
Bibliography 97
Index of defined terms 98

Figure 201.101 – MEDICAL BED, general arrangement (example, schematic presentation only) 13
Figure 201.102 – HEAD DOWN TILT, Example 15
Figure 201.103 – FOOT DOWN TILT, Example 15
Figure 201.104 – Entrapment test tools 17
Figure 201.105 – Loading pad 18
Figure 201.106 – Impactor 18
Figure 201.107 – Ball chain loop and spherical mass 19
Figure 201.108 – Graphic symbol for maximum PATIENT weight and SAFE WORKING LOAD 20

Figure 201.109 – Graphic symbol for mass; weight 20
Figure 201.110 – Graphic symbol for machine washable MEDICAL BED 20
Figure 201.111 – Graphic symbol for jet stream washable MEDICAL BED 21
Figure 201.112 – Graphic symbol for manual cleaning only 21
Figure 201.113 – Physical description of PATIENT 22
Figure 201.114 – MEDICAL BED function controls or actuators or both: guidelines for creating graphic symbols 23
Figure 201.115 – Example of MEDICAL BED with segmented or split SIDE RAIL 27
Figure 201.116 – Example of MEDICAL BED with single piece SIDE RAIL 28
Figure 201.117 – Allowable spacing for fingers in areas of normal reach around the perimeter of the MATTRESS SUPPORT PLATFORM 33
Figure 201.118 – Example using barriers for clearance measurement around the perimeter of the MATTRESS SUPPORT PLATFORM to mitigate PATIENT finger entrapment 34
Figure 201.119 – Clearance areas 35
Figure 201.120 – Retention of loop and mass 37
Figure 201.121 – Lateral stability test along the side of the MEDICAL BED 39
Figure 201.122 – Longitudinal stability test with removable FOOT BOARD 40
Figure 201.123 – Longitudinal stability test with fixed HEAD BOARD and FOOT BOARD 40
Figure 201.124 – Distribution of SAFE WORKING LOAD / maximum PATIENT weight for tests 44
Figure 201.125 – Positions of dynamic loading (see Figure 201.103 for loading pad and Figure 201.104 for impactor) 47
Figure 201.126 – Application of forces for test of SIDE RAIL 49
Figure 201.127 – Height of SIDE RAIL 52
Figure 201.128 – Direction of movement for rough handling test 59
Figure 201.129 – Configurations of the MATTRESS SUPPORT PLATFORM 61
Figure AA.1 – Example of marking to select recommended mattresses specified by the MANUFACTURER 68
Figure AA.2 – Example of marking for detachable SIDE RAILS specified by the MANUFACTURER 68
Figure AA.3 – Examples of marking on the MEDICAL BED of storage location for wired and wireless PENDANT CONTROLS 69
Figure AA.4 – Resultant forces without mattress 72
Figure AA.5 – Resultant forces with mattress 72
Figure AA.6 – Example of 60 mm gap measurement of B 73
Figure AA.7 – Angle measurement example of B 73
Figure AA.8 – Placement of measurement TOOL for measurement of D 74
Figure AA.9 – Example of area D measurement that passes 74
Figure AA.10 – Example of area D measurement that fails (on limit) 75
Figure AA.11 – Example of area D measurement that fails 75
Figure AA.12 – Example of potential PATIENT entrapment in area A within the SIDE RAIL 75
Figure AA.13 – Example of potential PATIENT entrapment in area A below the SIDE RAIL 76
Figure AA.14 – Example of potential PATIENT entrapment in area B 76
Figure AA.15 – Example of potential PATIENT entrapment in area C between split SIDE RAIL 76

Figure AA.16 – Example of potential PATIENT entrapment in area C between SIDE RAIL and HEAD BOARD 76
Figure AA.17 – Example of potential PATIENT entrapment in area D 77
Figure AA.18 – Example of potential PATIENT entrapment in area A below a single piece SIDE RAIL 77
Figure AA.19 – Concept of potential PATIENT foot-first entrapment tool for area A below a SIDE RAIL 78
Figure BB.1 – Schematic presentation of under MEDICAL BED clearance 86
Figure BB.2 – Recommendations and requirements regarding angles for different sections of the MATTRESS SUPPORT PLATFORM 88
Figure CC.1 – Wedge tool 90
Figure CC.2 – V-shaped opening in relation to B 91
Figure CC.3 – Acceptance criteria in relation to area B 92
Figure CC.4 – Positioning of wedge TOOL 93
Figure CC.5 – Acceptance criteria in relation to area C between HEAD BOARD and FOOT BOARD 93
Figure CC.6 – Acceptance criteria in relation to area C between split SIDE RAILS 94

Table 201.101 – Protection against PATIENT entrapment 29
Table 201.102 – Minimum SAFE WORKING LOADS 43
Table 201.103 – Protection against inadvertent PATIENT falls 53
Table 24 – Allowable maximum temperatures for skin contact with MEDICAL BED APPLIED PARTS 53
Table 201.104 – Machine washable compliance PROCEDURE 55