

ISO 389-5:2006-11 (E)

**Acoustics - Reference zero for the calibration of audiometric equipment - Part 5:
Reference equivalent threshold sound pressure levels for pure tones in the
frequency range 8 kHz to 16 kHz**

Contents		Page
Foreword		iv
Introduction		v
1 Scope		1
2 Normative references		1
3 Terms and definitions		1
4 Specifications		2
Annex A (informative) Notes on the derivation of the reference equivalent threshold sound pressure levels for audiometric earphones in the frequency range from 8 kHz to 16 kHz		3
Annex B (informative) RETSPL values for Koss HV/1A		4
Bibliography		6