ISO 8655-7:2022-04 (E)

Piston-operated volumetric apparatus - Part 7: Alternative measurement procedures for the determination of volume

Contents

Page

Foreword v

Introduction vi

1 Scope 1

2 Normative references 1

3 Termsanddefinitions 2

4 General requirements 2
4.1 Metrological confirmation 2
4.2 Uncertainty of measurement 2
4.3 Operator qualification 2

5 Performance requirements 3
5.1 Performance tolerances 3
5.2 Operator impact 3

6 Test conditions 3
6.1 General 3
6.2 Test equipment 3
6.3 Test room, environmental conditions 3
6.4 Test volumes 4
6.4.1 Fixed volume POVA 4
6.4.2 Adjustable volume POVA 4
6.5 Number of measurements per test volume 4
6.6 Test liquids 5

7 Evaluation 5
7.1 Mean volume 5
7.2 Systematic error of measurement 5
7.3 Random error of measurement 6

8 Test methods 6
8.1 General 6
8.2 Gravimetric method 7
8.3 Dual-dye ratiometric photometric method 7
8.4 Single dye photometric method 8
8.5 Hybrid photometric/gravimetric method for multichannel POVA 8
8.6 Titration method 8
8.7 Batch testing 8

9 Dispense procedures 8
9.1 General 8
9.2 Preparation 9
9.3 Single-channel air displacement pipettes (in accordance with ISO 8655-2) 9
9.3.1 General 9
9.3.2 Test cycle 9
9.4 Multi-channel pipettes (in accordance with ISO 8655-2) 10
9.5 Positive displacement pipettes (in accordance with ISO 8655-2) 11

9.6 Burettes (in accordance with ISO 8655-3) 11
9.7 Dilutors (in accordance with ISO 8655-4) 12
9.7.1 General 12
9.7.2 Test cycle 12
9.8 Dispensers (in accordance with ISO 8655-5) 13
9.9 Syringes (in accordance with ISO 8655-9) 13
9.9.1 General 13
9.9.2 Test cycle 13

10 Reporting of results 14

ISO 8655-7:2022(E) Annex A (normative) Gravimetric procedure 16

Annex B (normative) Dual-dye ratiometric photometric procedure 21

Annex C (normative) Single dye photometric procedure 29

Annex D (normative) Photometric/gravimetric hybrid procedure 33

Annex E (normative) Titrimetric procedure 41

Annex F (normative) Conversion of liquid mass to volume 45

Bibliography 48