

ISO/TR 16153:2023-02 (E)

Determination of uncertainty for volume measurements of a piston-operated volumetric apparatus using a photometric method

Contents		Page
Foreword		iv
Introduction		v
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	Modelling the measurement	1
5	General procedure for the uncertainty calculation	4
6	Standard uncertainty components associated with the measuring system (photometric measurement procedure)	4
6.1	General information on the estimation of standard uncertainty components	4
6.2	Standard uncertainty of the copper(II) chloride solution volume	5
6.3	Standard uncertainty of the cuvette mixture absorbance at 520 nm	5
6.4	Standard uncertainty of the cuvette starting absorbance at 730 nm	6
6.5	Standard uncertainty of the cuvette starting absorbance at 520 nm	6
6.6	Standard uncertainty of the volume of Ponceau S and copper(II) chloride solutions used in calibrators	7
6.7	Standard uncertainty of the absorbances of the calibrator solutions	7
7	Standard uncertainty components associated with the POVA	9
7.1	Standard uncertainty of the resolution	9
7.2	Standard uncertainty of the setting	9
7.3	Standard uncertainty related to air cushion effects	9
7.4	Standard uncertainty of the cubic expansion coefficient	9
8	Standard uncertainty components associated with the liquid delivery process	10
8.1	Repeatability (experimental standard deviation)	10
8.2	Reproducibility	10
9	Combined standard uncertainty of measurement associated with the systematic error of mean volume	11
10	Sensitivity coefficients	12
11	Coverage factor k	13
12	Expanded uncertainty of measurement associated with the volume \bar{V}	14
13	Example for determining the uncertainty of the volume measurement of POVA	14
13.1	Measurement conditions	14
13.2	Results	16
13.2.1	Standard uncertainty of the POVA mean volume	16
13.2.2	Expanded uncertainty of the measurement	17
13.2.3	Result of measurement	17
13.2.4	Uncertainty in use and corrections for pressure changes	17
13.2.5	General remarks	17
13.2.6	Note on the conformity with ISO/IEC Guide 98-3	17
Annex A (informative)	Approaches for the estimation of uncertainty in use of a single delivered volume	19
Annex B (informative)	Volume correction due to pressure changes	22
Bibliography		24