

DIN EN 15703-2:2015-03 (E)

Copper and copper alloys - Determination of manganese content - Part 2: Flame atomic absorption spectrometric method (FAAS)

Contents		Page
Foreword		3
1	Scope	4
2	Normative references	4
3	Principle	4
4	Reagents	4
4.1	Hydrochloric acid, HCl (= 1,19 g/ml)	4
4.2	Nitric acid, HNO ₃ (= 1,40 g/ml)	4
4.3	Hydrofluoric acid, HF (= 1,13 g/ml)	4
4.4	Nitric acid solution, 1 + 1	4
4.5	Lanthanum(III) chloride solution, 100 g/l	4
4.6	Sulphuric acid, H ₂ SO ₄ (= 1,84 g/ml)	4
4.7	Sulphuric acid solution, 1 + 9	5
4.8	Manganese stock solution, 1,0 g/l Mn	5
4.9	Manganese standard solution, 0,10 g/l	5
4.10	Manganese standard solution, 0,010 g/l	5
4.11	Copper matrix solution, 20 g/l Cu	5
4.12	Copper matrix solution, 2 g/l Cu	5
5	Apparatus	5
5.1	Atomic absorption spectrometer, fitted with an air/acetylene burner	5
5.2	Manganese hollow-cathode lamp	6
6	Sampling	6
7	Procedure	6
7.1	Preparation of the test portion solution	6
7.2	Blank test	7
7.3	Check test	7
7.4	Establishment of the calibration curve	7
7.5	Determination	10
8	Expression of results	10
8.1	Use of the calibration curve	10
8.2	Use of the bracketing method	11
9	Precision	12
10	Test report	13
Bibliography		14