

# ISO 16000-42:2023-08 (E)

## Indoor air - Part 42: Measurement of the particle number concentration by condensation particle counters

---

<b>Contents</b>		<b>Page</b>
Foreword .....		v
Introduction .....		vi
1	Scope .....	1
2	Normative references .....	1
3	Terms and definitions .....	1
4	Abbreviated terms .....	3
5	Sources of airborne particles .....	4
5.1	General .....	4
5.2	Combustion of organic material .....	5
5.3	Smoking .....	5
5.4	Cooking .....	5
5.5	Particle formation -- Formation of secondary organic aerosol .....	5
5.6	Outdoor air .....	5
5.7	Other sources .....	5
6	Dynamicsof ultrafine particles indoors .....	6
6.1	General .....	6
6.2	Infiltration and exfiltration .....	7
6.3	Deposition .....	7
6.4	Particle formation, phase transition and coagulation .....	7
7	Principle of measurement .....	8
7.1	General .....	8
7.2	Working fluid .....	8
7.3	Minimal detection size .....	10
7.3.1	General .....	10
7.3.2	Optical detection after enlargement .....	10
7.3.3	Particle size distribution .....	11
7.4	CPC minimal requirement .....	11
7.5	General sampling requirements and recommendations .....	13
8	Measurement strategy .....	13
8.1	General .....	13
8.2	Average room concentration .....	14
8.2.1	General .....	14
8.2.2	Resting state without activity .....	15
8.2.3	Resting state with equipment activity .....	15
8.2.4	Active state .....	15
8.3	Source investigation/identification .....	15
8.4	Infiltration from outdoor or connecting rooms .....	16
8.5	Measurement in vehicle cabins .....	17
8.6	Success of control and mitigation measures .....	17
9	Quality assurance and uncertainty evaluation .....	17
9.1	General .....	17

9.2	Instrument parameters .....	18
9.3	CPC's settings check .....	18
9.4	Performance check, zero check or leak check .....	18
9.5	Uncertainty .....	19
10	Evaluation and reporting of the results .....	19
Annex A (informative)	Examples of particle number concentrations encountered during room user activities .....	21
Annex B (informative)	Determination of the particle number size distribution of indoor aerosol using a differential mobility aerosol spectrometer .....	22
Annex C (informative)	Water-CPCs .....	25
Annex D (informative)	Checklist to collect information useful for interpreting indoor measurement of particle number concentration .....	27
Bibliography	.....	31