

DIN EN 14626:2026-10 (E)

Ambient air - Standard method for the measurement of the concentration of carbon monoxide by non-dispersive infrared spectroscopy

Contents

Page

European foreword	5
1 Scope	6
2 Normative references	7
3 Terms and definitions	8
4 Abbreviated terms	13
5 Principle	13
5.1 General	13
5.2 Measuring principle	13
5.3 Type testing	14
5.4 Field operation and quality control	14
6 Sampling	15
6.1 General	15
6.2 Sampling location	15
6.3 Sampling system	15
6.3.1 Construction	15
6.3.2 Particle filter	16
6.3.3 Loss of carbon monoxide	16
6.3.4 Conditioning	16
6.4 Control and regulation of sample flow rate	16
6.5 Sampling pump for the manifold	17
7 Analyser equipment	17
7.1 General	17
7.2 Details about analyser equipment	17
7.3 Pressure measurement	17
7.4 Flow rate indicator	17
7.5 Sampling pump for the analyser	18
7.6 Particle filter	18
8 Type testing of carbon monoxide analysers	18
8.1 General	18
8.2 Relevant performance characteristics and performance criteria	19
8.3 Design change	23
8.4 Procedures for determination of the performance characteristics during the laboratory test	23
8.4.1 General	23
8.4.2 Test conditions	23
8.4.3 Response time	26
8.4.4 Short-term drift	28
8.4.5 Repeatability standard deviation	28
8.4.6 Lack of fit of linearity of the calibration function	29
8.4.7 Sensitivity coefficient to sample gas pressure	30
8.4.8 Sensitivity coefficient to the sample gas temperature	31
8.4.9 Sensitivity coefficient to the surrounding temperature	31
8.4.10 Sensitivity coefficient to electrical voltage	32

8.4.11	Interferents	33
8.4.12	Averaging test	34
8.4.13	Difference sample/calibration port	35
8.5	Determination of the performance characteristics during the field test	36
8.5.1	General	36
8.5.2	Selection of a monitoring station for the field test	36
8.5.3	Operational requirements	37
8.5.4	Long-term drift	38
8.5.5	Reproducibility standard deviation under field conditions	39
8.5.6	Period of unattended operation	39
8.5.7	Period of availability of the analyser	39
8.6	Expanded uncertainty calculation for type testing	40
9	Field operation and ongoing quality control	41
9.1	General	41
9.2	Suitability evaluation	41
9.2.1	General	41
9.2.2	Analyser for a monitoring station or task	41
9.3	Initial installation	43
9.4	Ongoing quality assurance/quality control	43
9.4.1	General	43
9.4.2	Frequency of calibrations, checks and maintenance	44
9.5	Calibration of the analyser	47
9.5.1	General	47
9.5.2	Calibration gases	48
9.5.3	Data adjustment function	48
9.5.4	Testing the sampling system	48
9.5.5	Treatment of data after exceedance of performance criteria	50
9.6	Checks	52
9.6.1	Zero and span checks	52
9.6.2	Lack-of-fit	53
9.6.3	Testing the sampling system	54
9.7	Maintenance	56
9.7.1	Change of particle filters	56
9.7.2	Maintenance of sampling system	56
9.7.3	Change of consumables as applicable	56
9.7.4	Preventive/routine maintenance of components of the analyser	56
9.8	Data handling and data reports	56
9.9	Measurement uncertainty	57
10	Expression of results	57
11	Test reports and documentation	58
11.1	Type testing	58
11.2	Field operation	59
11.2.1	Suitability evaluation	59
11.2.2	Documentation	59
11.2.3	Ambient air quality data reports	60
Annex A (normative)	Test of lack-of-fit	61
A.1	Establishment of the regression line	61
A.2	Calculation of the residuals of the averages	61
Annex B (informative)	Sampling equipment	63
Annex C (informative)	Schematics of non-dispersive infrared spectrometer	65
Annex D (informative)	Manifold testing	67
D.1	Procedure for applying test gas	67
D.2	Procedure for the cross test	68

D.2.1	General	68
D.2.2	Initial stage	68
D.2.3	Stage 1	69
D.2.4	Stage 2	69
D.2.5	Data Processing	69
D.2.6	Evaluation	69
Annex E (normative) Type testing		71
E.1	Type testing and uncertainty calculation	71
E.1.1	Type testing	71
E.1.2	Uncertainty calculation	71
E.2	Type testing Requirement a)	71
E.3	Type testing Requirement b)	73
E.3.1	General	73
E.3.2	Calculation of standard uncertainties	76
E.3.3	Example calculation	83
E.4	Type testing Requirement c)	85
E.5	Type testing Requirement d)	85
E.5.1	General	85
E.5.2	Combined standard uncertainty	87
E.5.3	Absolute expanded uncertainty	87
E.5.4	Relative expanded uncertainty	88
E.5.5	Calculation of standard uncertainties	88
E.5.6	Example calculation	90
Annex F (informative) Calculation of uncertainty in field operation at the 8-hour limit value		92
F.1	General	92
F.2	Combined standard uncertainty	92
F.3	Standard uncertainties	93
F.3.1	General	93
F.3.2	Influence quantities	94
F.3.3	Interferents	97
F.3.4	Averaging effect	98
F.3.5	Reproducibility under field conditions	98
F.3.6	Long-term drift at zero	98
F.3.7	Long-term drift at level of the 8-hour limit value	98
F.3.8	Zero gas	99
F.3.9	Calibration gas	99
F.3.10	Difference sample/calibration port	99
F.4	Example calculation	100
Annex G (informative) Test stand for the test point "sensitivity coefficient of sample gas pressure"		102
Annex H (informative) Significant changes		104
Bibliography		105