# ISO 23496:2019 (E)

## Determination of pH value — Reference buffer solutions for the calibration of pH measuring equipment

# Contents


Foreword

1 Scope

2 Normative references

3 Terms and definitions

4 Basic information

- 4.1 General
- 4.2 pH values of primary reference buffer solutions
- 4.3 pH values of secondary reference buffer solutions

5 Preparation of primary and secondary reference buffer solutions

- 5.1 General
- 5.2 Reference buffer solution A, pH(R) = 1,6791 1 The last digit is not significant and therefore printed as subscript.(25 °C), Potassium tetraoxalate dihydrate
- 5.3 Reference buffer solution B, pH(R) = 3,557 (25 °C), Potassium hydrogen tartrate
- 5.4 Reference buffer solution C, pH(R) = 4,005 (25 °C), Potassium hydrogen phthalate
- 5.5 Reference buffer solution D, pH(R) = 6,865 (25 °C), Phosphate
- 5.6 Reference buffer solution E, pH(R) = 7,413 (25 °C), Phosphate
- 5.7 Reference buffer solution F, pH(R) = 9,180 (25 °C), Borax
- 5.8 Reference buffer solution G, pH(R) = 12,454 (25 °C), Calcium hydroxide
- 5.9 Reference buffer solution H, pH(R) = 3,776 (25 °C), Potassium dihydrogen citrate
- 5.10 Reference buffer solution I, pH(R) = 10,012 (25 °C), Sodium carbonate/sodium hydrogen carbonate

6 Storage and shelf life

7 Examples of pH(R) values of reference buffer solutions

8 Additional properties of reference buffer solutions

Annex A (informative) Examples of pH(R) values as a function of temperature at temperatures from 55 °C to 95 °C

Annex B (informative) Volumetric method

- B.1 General
- B.2 Reference buffer solution A, pH(R) = 1,679 (25 °C), Potassium tetraoxalate dihydrate
- B.3 Reference buffer solution B, pH(R) = 3,557 (25 °C), Potassium hydrogen tartrate
- B.4 Reference buffer solution C, pH(R) = 4,005 (25 °C), Potassium hydrogen phthalate
- B.5 Reference buffer solution D, pH(R) = 6,865 (25 °C), Phosphate
- B.6 Reference buffer solution E, pH(R) = 7,413 (25 °C), Phosphate
- B.7 Reference buffer solution F, pH(R) = 9,180 (25 °C), Borax
- B.8 Reference buffer solution G, pH(R) = 12,454 (25 °C), Calcium hydroxide
- B.9 Reference buffer solution H, pH(R) = 3,776 (25 °C), Potassium dihydrogen citrate
- B.10 Reference buffer solution I, pH(R) = 10,012 (25 °C), Sodium carbonate/sodium hydrogen carbonate


Page count: 13