DIN EN ISO 10650:2026-12 (E)

Dentistry - Powered polymerization activators (ISO 10650:2026)

Contents

Page

Foreword ..... **iv**

Introduction ..... **v**

1 Scope **1**

2 Normative references **1**

3 Terms and definitions **1**

4 Classification **2**

5 Requirements **3**

5.1 General 3

5.1.1 Design 3

5.1.2 Connection 3

5.1.3 Operating controls 3

5.2 Radiant exitance 3

5.2.1 General 3

5.2.2 Radiant exitance in the 380 nm to 515 nm wavelength region 3

5.2.3 Radiant exitance in the wavelength region below 380 nm 3

5.2.4 Radiant exitance in the wavelength region above 515 nm 3

5.3 Instructions for use 4

5.4 Marking 4

5.5 Packaging 4

6 Sampling **4**

7 Measurement and test methods **4**

7.1 General 4

7.1.1 General provisions for tests 4

7.1.2 Atmospheric conditions 4

7.2 Inspections 4

7.2.1 Visual inspection of powered polymerization activator 4

7.2.2 Manual inspection of powered polymerization activator 5

7.2.3 Visual inspection of manufacturer's document 5

7.3 Measurement of the optical cross-sectional area of the light guide/light-emitting tip 5

7.3.1 Apparatus 5

7.3.2 Procedure 5

7.4 Measurement of radiant exitance 5

7.4.1 Method A using spectrometer 5

7.4.2 Method B using filters 8

8 Information (instructions for use) to be supplied by the manufacturer **12**

9 Marking **13**

10 Packaging **13**

Bibliography **14**