

DIN EN ISO 11979-3:2013-04 (E)

Ophthalmic implants - Intraocular lenses - Part 3: Mechanical properties and test methods (ISO 11979-3:2012)

Contents	Page
Foreword.....	3
Introduction.....	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Requirements	5
4.1 General.....	5
4.2 Tolerances and dimensions.....	6
4.3 Clearance analysis (anterior chamber lenses only).....	6
4.4 Compression force.....	6
4.5 Axial displacement in compression.....	6
4.6 Optic decentration.....	7
4.7 Optic tilt.....	7
4.8 Angle of contact.....	7
4.9 Compression force decay.....	7
4.10 Dynamic fatigue durability.....	7
4.11 Surgical manipulation.....	8
4.12 Surface and bulk homogeneity.....	8
5 Recovery of properties following simulated surgical manipulation	8
6 Additions for accommodating IOLs (AIOLs)	8
Annex A (normative) Measurement of compression force	10
Annex B (normative) Measurement of axial displacement in compression	13
Annex C (normative) Measurement of optic decentration	16
Annex D (normative) Measurement of optic tilt	19
Annex E (normative) Measurement of angle of contact	23
Annex F (normative) Testing of compression force decay	26
Annex G (normative) Testing of dynamic fatigue durability	27
Annex H (informative) Measurement of loop pull strength	29
Annex I (informative) Clearance analysis	31
Annex J (informative) Precision	34
Bibliography	35