

# ISO 5840-1:2021 (E)

## Cardiovascular implants — Cardiac valve prostheses — Part 1: General requirements

---

### Contents

	Foreword
	Introduction
1	Scope
2	Normative references
3	Terms and definitions
4	Abbreviations
5	Fundamental requirements
6	Device description
6.1	General
6.2	Intended use
6.3	Design inputs
6.3.1	Operational specifications
6.3.2	Performance specifications
6.3.3	Implant procedure
6.3.4	Packaging, labelling, and sterilization
6.4	Design outputs
6.5	Design transfer (manufacturing verification/validation)
6.6	Risk management
7	Design verification and validation
7.1	General requirements
7.2	In vitro assessment
7.2.1	General
7.2.2	Test conditions, sample selection and reporting requirements
7.2.2.1	Test articles and sample selection
7.2.2.2	Test conditions
7.2.2.3	Reporting requirements
7.2.3	Material property assessment
7.2.3.1	General
7.2.3.2	Biological safety
7.2.3.3	Material and mechanical property testing
7.2.4	Hydrodynamic performance assessment
7.2.5	Structural performance assessment
7.2.5.1	General
7.2.5.2	Implant durability assessment
7.2.5.3	Device structural component fatigue assessment
7.2.5.4	Component corrosion assessment
7.2.6	Design- or procedure-specific testing
7.2.7	Device MRI compatibility
7.2.8	Simulated use
7.2.9	Human factors/usability assessment
7.2.10	Implant thrombogenic and haemolytic potential assessment
7.3	Preclinical in vivo evaluation
7.4	Clinical investigations
Annex A	(informative) Rationale for the provisions of ISO 5840-1

- A.1 Rationale for a risk-based approach
- A.2 Rationale for preclinical in vivo evaluation
- A.3 Rationale for design verification and design validation testing
- A.4 Rationale for echocardiographic assessment
- A.5 Rationale for clinical evaluation reporting
- A.6 Rationale for device sizing within labelling and instructions for use
- A.7 Rationale for human factors engineering

**Annex B (normative) Packaging**

- B.1 Requirements
- B.2 Principle
- B.3 Containers
  - B.3.1 Unit container(s)
  - B.3.2 Outer container

**Annex C (normative) Product labels, instructions for use, and training**

- C.1 General
  - C.1.1 General requirements
  - C.1.2 Unit-container label
  - C.1.3 Outer-container label
  - C.1.4 Instructions for use
  - C.1.5 Labels for medical records
- C.2 Training for physicians and support staff

**Annex D (normative) Sterilization**

**Annex E (normative) In vitro test guidelines for paediatric devices**

- E.1 General and paediatric definitions
- E.2 Pulsatile flow test conditions — Left side
- E.3 Pulsatile flow test conditions — Right side
- E.4 Steady back pressure and forward flow conditions — Left side
- E.5 Steady back pressure and forward flow conditions — Right side
- E.6 Accelerated wear testing (AWT) test conditions — Left side
- E.7 Accelerated wear testing (AWT) test conditions — Right side
- E.8 FEA/life analysis conditions — Left side
- E.9 FEA/life analysis conditions — Right side

**Annex F (informative) Corrosion assessment**

- F.1 Rationale
- F.2 General
- F.3 Pitting corrosion
- F.4 Crevice corrosion
- F.5 Galvanic corrosion
- F.6 Corrosion fatigue
- F.7 Fretting (wear) and fretting corrosion
- F.8 Post-fatigue corrosion evaluation

**Annex G (informative) Echocardiographic protocol**

- G.1 General
- G.2 Echocardiographic studies
- G.3 Data collected

**Annex H (informative) Assessment of implant thrombogenic and haemolytic potential**

- H.1 Rationale
- H.2 General
- H.3 Experimental flow field assessment
  - H.3.1 General
  - H.3.2 Test apparatus requirements
  - H.3.3 Test procedure
  - H.3.4 Test report
- H.4 Computational flow field assessment
  - H.4.1 General
  - H.4.2 Computational Model
  - H.4.3 Error analysis and estimation

- H.4.4 Computational simulations
- H.4.5 Study report
- H.5 Ex vivo blood testing
- H.5.1 General
- H.5.2 Test apparatus requirements
- H.5.3 Test procedure
- H.5.4 Test report

**Annex I (informative) Guidelines for hydrodynamic performance characterization by steady flow testing**

- I.1 General
- I.2 Steady forward flow testing
  - I.2.1 Measuring equipment accuracy
  - I.2.2 Test apparatus requirements
  - I.2.3 Test procedure
  - I.2.4 Test report
- I.3 Steady back flow leakage testing
  - I.3.1 Measuring equipment accuracy
  - I.3.2 Test apparatus requirements
  - I.3.3 Test procedure
  - I.3.4 Test report

**Annex J (normative) Durability testing**

- J.1 Rationale
- J.2 General
- J.3 Accelerated wear testing
  - J.3.1 General
  - J.3.2 Sample requirements
  - J.3.3 Test apparatus requirements
  - J.3.4 Test procedure
  - J.3.5 Test report
- J.4 Dynamic failure mode testing
  - J.4.1 General
  - J.4.2 Sample requirements
  - J.4.3 Test apparatus requirements
  - J.4.4 Test procedure
  - J.4.5 Test report
- J.5 Real-time wear testing
  - J.5.1 General
  - J.5.2 Test apparatus requirements
  - J.5.3 Test apparatus requirements
  - J.5.4 Test procedure
  - J.5.5 Test report

**Annex K (informative) Fatigue assessment**

- K.1 General
- K.2 Determination of in vivo boundary conditions
- K.3 Structural component stress/strain analysis
- K.4 Material fatigue strength determination
- K.5 Fatigue safety factor or probability of fatigue fracture determination
- K.6 Component fatigue demonstration test

**Annex L (normative) Clinical investigation endpoints for heart valve replacement devices**

- L.1 General
- L.2 Single endpoints
  - L.2.1 General
  - L.2.2 Safety
  - L.2.3 Effectiveness
- L.3 Heart failure hospitalization definition
- L.4 Composite endpoints
- L.5 Timing of endpoints